



Brazilian Journal of Cardiovascular Surgery

ISSN: 0102-7638

ISSN: 1678-9741

Sociedade Brasileira de Cirurgia Cardiovascular

Xing, Zhenhua; Pei, Junyu; Huang, Jiabing; Hu, Xinqun; Gao, Shan
Diagnostic Performance of QFR for the Evaluation of Intermediate
Coronary Artery Stenosis Confirmed by Fractional Flow Reserve
Brazilian Journal of Cardiovascular Surgery, vol. 34, no. 2, 2019, March-April, pp. 165-172
Sociedade Brasileira de Cirurgia Cardiovascular

DOI: 10.21470/1678-9741-2018-0234

Available in: <http://www.redalyc.org/articulo.oa?id=398959097009>

- How to cite
- Complete issue
- More information about this article
- Journal's webpage in redalyc.org

UAEM redalyc.org

Scientific Information System Redalyc

Network of Scientific Journals from Latin America and the Caribbean, Spain and Portugal

Project academic non-profit, developed under the open access initiative

Diagnostic Performance of QFR for the Evaluation of Intermediate Coronary Artery Stenosis Confirmed by Fractional Flow Reserve

Zhenhua Xing¹, MD; Junyu Pei¹, MD; Jiabing Huang¹, MD; Xinqun Hu¹, MD; Shan Gao², MD

DOI: 10.21470/1678-9741-2018-0234

Abstract

Introduction: Quantitative flow ratio (QFR) is a novel method enabling efficient computation of FFR from three-dimensional quantitative coronary angiography (3D QCA) and thrombolysis in myocardial infarction (TIMI) frame counting. We decided to perform a systematic review and quantitative meta-analysis of the literature to determine the correlation between the diagnosis of functionally significant stenosis obtained by QFR versus FFR and to determine the diagnostic accuracy of QFR for intermediate coronary artery stenosis.

Methods: We searched PubMed, Embase, and Web of Science for studies concerning the diagnostic performance of QFR. Our meta-analysis was performed using the DerSimonian and Laird random effects model to determine sensitivity, specificity, positive likelihood ratio (LR+), negative likelihood ratio (LR-), and

diagnostic odds ratio (DOR). The sROC was used to determine diagnostic test accuracy.

Results: Nine studies consisting of 1175 vessels in 1047 patients were included in our study. The pooled sensitivity, specificity, LR+, LR-, and DOR for QFR were 0.89 (95% CI: 0.86-0.92), 0.88 (95% CI: 0.86-0.91), 6.86 (95% CI: 5.22-9.02), 0.14 (95% CI: 0.10-0.21), and 53.05 (95% CI: 29.75-94.58), respectively. The area under the summary receiver operating characteristic (sROC) curve for QFR was 0.94.

Conclusion: QFR is a simple, useful, and noninvasive modality for diagnosis of functional significance of intermediate coronary artery stenosis.

Keywords: Myocardial Fractional Flow Reserve. Coronary Artery Disease. Quantitative Flow Ratio. Sensitivity and Specificity. Meta-Analysis [Publication Type].

Abbreviations, acronyms & symbols			
3D	= Three-dimensional	LR+	= Positive likelihood ratio
3D QCA	= Three-dimensional quantitative coronary angiography	LR-	= Negative likelihood ratio
AUC	= Area under the curve	OR	= Odds ratio
CI	= Confidence interval	MI	= Myocardial infarction
DOR	= Diagnostic odds ratio	PCI	= Percutaneous coronary intervention
EAPCI	= European Association of Percutaneous Cardiovascular Interventions	PRISMA	= Preferred Reporting Items for Systematic Review and Meta-Analysis Protocols
ESC	= European Society of Cardiology	QFR	= Quantitative flow ratio
FFR	= Fractional flow reserve	QUADAS-2	= Quality Assessment of Diagnostic Accuracy Studies 2
FP	= False positive	sROC	= Summary receiver operating characteristic
FAVOR	= Functional Diagnostic Accuracy of Quantitative Flow Ratio in Online Assessment of Coronary Stenosis study	STEMI	= ST-elevation myocardial infarction
FN	= False negative	TIMI	= Thrombolysis in myocardial infarction
LR	= Likelihood ratio	TN	= True negative
		TP	= True positive

¹Department of Cardiology, Second Xiangya Hospital, Changsha, China.

²Department of Geriatrics, Second Xiangya Hospital, Changsha, China.

This study was carried out at the Department of Geriatrics and the Department of Cardiovascular Medicine, Second Xiangya Hospital, Central South University, Changsha, China.

Correspondence Address:

Shan Gao

 <http://orcid.org/0000-0001-8969-494X>

Second Xiangya Hospital, Central South University

No.139 Renmin Road, Changsha, Hunan 410011, China

E-mail: xing2012x@csu.edu.cn

Article received on August 8th, 2018.

Article accepted on October 17th, 2018.

INTRODUCTION

Accurate evaluation of coronary artery disease, especially intermediate coronary artery stenosis, is crucial for the evaluation of myocardial ischemia and next treatment. The gold standard for diagnosis and confirmation of functional significance of a stenosis is the fractional flow reserve (FFR). Previous studies have demonstrated that FFR-guided coronary revascularization increases the ratio of event-free survival when compared with a coronary stenosis-guided strategy^[1,2]. Despite these advantages, the clinical application of FFR has been variable and slow^[3]. FFR requires not only the hyperemic state, but also additional cost, time, and efforts.

Quantitative flow ratio (QFR) is a novel method enabling efficient computation of FFR from three-dimensional quantitative coronary angiography (3D QCA) and thrombolysis in myocardial infarction (TIMI) frame counting^[4]. Compared with FFR, QFR does not require any invasive physiological measurements, pharmacological hyperemia induction, and additional cost. The recently FAVOR (Functional Diagnostic Accuracy of Quantitative Flow Ratio in Online Assessment of Coronary Stenosis) II China study showed solid results for QFR computation in identifying the presence of functionally significant stenosis in eligible patients^[5]. Several studies have been published in the literature addressing the correlation between the assessment of functionally significant stenosis obtained by QFR versus FFR and addressing the diagnostic accuracy of QFR for intermediate coronary artery stenosis^[5,6]. The purpose of our study was to perform a systematic review and quantitative meta-analysis of the literature to determine the correlation between the diagnosis of functionally significant stenosis obtained by QFR versus FFR, and to determine the diagnostic accuracy of QFR for intermediate coronary artery stenosis.

METHODS

This protocol is reported following the Preferred Reporting Items for Systematic Review and Meta-Analysis Protocols (PRISMA) guidelines^[7]. We searched PubMed, Embase, and Web of Science published before March 15, 2018. The keywords used for search were "QFR or Quantitative flow ratio". Results were limited to trials published in English. We manually searched reference lists of relevant studies and reviews, editorials, and letters to identify further articles. We used Endnote (Thompson ISI ResearchSoft, Philadelphia, USA) to manage relevant articles and remove duplicated articles.

Study Eligibility

The inclusion criteria for studies in the analysis were as follows: 1) The design was a diagnostic accuracy study; 2) The study assess the diagnostic performance of QFR compared with invasive FFR as the standard procedure; 3) Data from true positive (TP), false positive (FP), true negative (TN), false negative (FN), sensitivity and specificity can be retrieved or calculated. When relevant data were missing, authors were contacted by e-mail, before excluding the study due to inaccessibility of data.

Data Collection and Quality Assessment

Relevant data were initially extracted by two independent reviewers (Zh Xing and Jy Pei). Disagreements were resolved by consensus or by a third investigator (XQ Hu). We abstracted the following data from the selected articles: first author, publication date, study design, patient demographics; FFR threshold used to describe ischemia; and the data of TP, FP, TN, and FN. When different flow models of QFR were performed, contrast-flow QFR was preferred. Contrast-flow QFR was more accurate for predicting FFR ≤ 0.80 as compared with fixed-flow QFR^[4]. Included studies were analyzed by the Quality Assessment of Diagnostic Accuracy Studies 2 (QUADAS-2)^[8].

Data Analysis

The inter-reviewer agreement regarding the quality assessment of included studies was assessed by the Cohen kappa test. Our meta-analysis was performed using the DerSimonian and Laird random effects model to determine sensitivity, specificity, positive likelihood ratio (LR+), negative likelihood ratio (LR-), and diagnostic odds ratio (DOR). The sROC was used to determine diagnostic test accuracy. An area under the curve (AUC) between 0.75 and 0.92 represented a high degree of diagnostic accuracy, and an AUC between 0.93 and 0.96 was considered more accurate. In order to assess heterogeneity among the studies, the^[2] statistic was used. For^[2], a value $>50\%$ was considered of severe heterogeneity. The Spearman correlation coefficient was calculated to evaluate diagnostic threshold variation among the included studies.

We also performed a meta-regression analysis to identify predefined potential sources of heterogeneity. All statistical analyses were completed using Meta-DiSc (version 1.4).

RESULTS

Study Selection and Characteristics

The flowchart of our search and selection process was presented in Figure 1. Our combined search strategy identified possible relevant studies. Nine studies were retrieved for a more detailed evaluation. Finally, 9 studies consisting of 1175 vessels in 1047 patients met our inclusion criteria^[4-6, 9-15]. Characteristics of included studies were shown in Table 1. Clinical heterogeneity was mostly due to different inclusion criteria. Mejia-Renteria^[9] and Emori^[13] included patients with myocardial infarction. Emori^[13] included two different populations: patients with previous myocardial infarction (MI) and patients without previous MI. Spitaleri^[12] included patients with ST-elevation myocardial infarction (STEMI) and multivessel disease. Four studies were performed in Japan, two in China, one each in Spain, Italy, and Netherlands. The mean (SD) age was 63.2 years, and 68.1% of the patients were male. The quality assessment of included studies according to QUADAS-2 was presented in Supplementary Figure 1. In general, there was low risk of bias and low concern regarding applicability of all included studies.

Diagnostic Accuracy of QFR

The results of the included study were presented in Table 2. The accuracy ranged from 80% in Kameyama^[11] and 94% in




Fig. 1 - Flowchart for the identification of studies.

Spitaleri, et al.^[12]. Sensitivity ranged between 74% in Tu, et al.^[4] and 100% in van Rosendaal, et al.^[10] and the specificity ranged from 79% in Emori, et al.^[15] to 97% in Spitaleri, et al.^[12]. The correlation between QFR and FFR ranged from $r = 0.69$ to $r = 0.94$.

In pooled data weighted by the number of vessels, QFR had a combined sensitivity and the specificity of QFR for diagnosis of

functional significance of a stenosis according to FFR were 0.89 (95% CI: 0.86-0.92) and 0.88 (95% CI: 0.86-0.91), respectively, using a random effects model (Figure 2). No heterogeneity was found for both sensitivity ($I^2=38.3\%$, $P=0.10$) and specificity ($I^2=24.1\%$, $P=0.22$). The pooled estimate of positive likelihood ratio (LR+) and negative likelihood ratio (LR-) was 6.86 (95% CI: 5.22-9.02)

Table 1. Characteristics of included studies.

Study	Design	Country	Mean age (years)	Males (%)	Patients' characteristics	Cutoff
Emori et al. ^[15]	Multicenter	Japan	-	-	Intermediate stenosis with FFR	0.8
Xu et al. ^[5]	Prospective, multicenter	China	61	73.7	Intermediate stenosis with FFR	0.8
Yazaki et al. ^[6]	Retrospective, single-center	Japan	72.5	29.6	Intermediate stenosis with FFR	0.8
Emori et al. ^[13] MI*	Retrospective, single-center	Japan	69	83	Patients with previous MI undergoing CAG and FFR	0.8
Emori et al. ^[13] non-MI*	Retrospective, single-center	Japan	70	54	Intermediate stenosis with FFR	0.8
Tu et al. ^[4]	Prospective, multicenter	China	65.8	61	Intermediate stenosis with FFR	0.8
Kameyama et al. ^[11]	Multicenter	Japan	-	-	ACS patients with CAG and FFR	0.8
van Rosendaal et al. ^[10]	Prospective, single-center	Netherlands	64	71	Intermediate stenosis with FFR	0.8
Mejia-Renteria et al. ^[9]	Multicenter	Spain	-	-	Patients with CAG and FFR	0.8
Spitaleri et al. ^[12]	Prospective, multicenter	Italy	62	28	STEMI patients with MVD and FFR	0.8
Spitaleri et al. ^[12]	Prospective, multicenter	Denmark	61	67	Intermediate stenosis with FFR	0.8

*Emori 2018 contained two groups: patients with previous myocardial infarction (MI) and patients with no previous MI.

ACS=acute coronary syndrome; CAG=coronary angiography; FFR=fractional flow reserve; MVD=multivessel disease; STEMI=ST-elevation myocardial infarction

Table 2. Results of included studies in these meta-analyses.

Study	Included vessels (n)	Sensitivity (%)	Specificity (%)	Accuracy (%)	Correlation (r)
Emori et al. ^[15]	73	82	79	81	0.69
Xu et al. ^[5]	328	94.6	91.7	92.7	0.86
Yazaki et al. ^[6]	151	88.7	89.1	88.7	0.84
Emori et al. ^[13] MI	75	92	82	87	0.88
Emori et al. ^[13] non-MI	75	95	88	92	0.94
Tu et al. ^[4]	84	74	91	86	0.77
Kameyama et al. ^[11]	25	80	80	80	0.63
van Rosendaal et al. ^[10]	15	100	79	80	0.78
Mejia-Renteria et al. ^[9]	300	88	86	87	-
Spitaleri et al. ^[12]	49	88	97	94	0.90

MI=myocardial infarction

and 0.14 (95% CI: 0.10-0.21) (Figure 3). For QFR, Spearman's correlation coefficients were 0.619 ($P=0.102$), indicating that the diagnostic threshold effect did not exist in QFR data. The area under the ROC curve was 0.94 (Figure 4) and the diagnostic OR was 53.05 (95% CI: 29.75-94.58) (Supplementary Figure 2).

Meta-regression Analysis and Subgroup Analysis

Meta-regression was performed using the potential sources of heterogeneity among studies (age, country, sex, different inclusion criteria). We found no factor effecting the diagnostic accuracy.

Emori, et al.^[13] and Mejia-Renteria, et al.^[9] included patients with previous MI that might affect the diagnostic accuracy. Exclusion of these two trials slightly improved the specificity (0.90, 0.87-0.92), but did not affect sensitivity.

DISCUSSION

In the present study, we performed a systematic review of the diagnostic performance of QFR for functional significance of intermediate coronary artery stenosis compared with invasive FFR. Data from 1175 vessels in 1047 patients showed the situation in which QFR is helpful for surgeons to determine whether stents




Fig. 2 - Forest plot of the sensitivity and specificity of included study, summary sensitivity and specificity and I2 statistic for heterogeneity.

should be implanted at no additional cost, time, and effort. Our study is the first systemic review and meta-analysis that evaluates the diagnostic accuracy of QFR for the assessment of functionally significant stenosis confirmed by FFR.

FFR-guided percutaneous coronary intervention (PCI) is associated with a better outcome compared with revascularization based on angiographic stenosis severity alone in patients with intermediate coronary artery stenosis^[2,16,17]. FFR has the highest recommendation (class I, level A) in the European Society of Cardiology guideline on myocardial revascularization^[18]. Although FFR has better outcomes in patients with intermediate coronary artery stenosis, the clinical application has been limited

due to the invasive procedure with a pressure wire, the cost of pressure wire, and the side effects associated with induction of hyperemia. QFR, an angiographic index of coronary stenosis severity based on 3D QCA and thrombolysis in myocardial infarction (TIMI) frame counting, estimates FFR without invasive procedure^[4]. FAVOR Pilot Study and FAVOR II China study have shown good agreement of QFR with invasive FFR^[4,5]. However, the sample size was too small. Our meta-analysis included^[9] studies of which the majority were performed in China and Japan from 2016 to 2018, with a total of 1175 vessels in 1047 patients, and demonstrated that the diagnostic accuracy of QFR for functionally significant stenosis confirmed by FFR was high, with




Fig. 3 - Forest plot of LR+ and LR- of included study, summary sensitivity and specificity, and I² statistic for heterogeneity.

a summary sensitivity and specificity of 0.89 (95% CI: 0.86-0.92) and 0.88 (95% CI: 0.86-0.91), respectively.

Our studies included different study populations (stable coronary artery disease, suspected coronary artery disease, STEMI, previous MI). However, our meta-regression showed that different study populations did not affect our diagnostic accuracy. QFR was more often used in patients with stable coronary artery disease. A recent study has found that QFR may be a safe and reliable tool to guide revascularization in patients with STEMI and multivessel disease. Furthermore, Spitaleri found that functional complete revascularization evaluated by QFR showed a good 5-year outcome^[12].

However, the diagnostic accuracy of QFR for assessing the functional severity of coronary stenosis might be affected in coronary arteries related to previous MI^[13]. Microcirculatory resistance may affect this phenomenon. Mejia-Renteria, et al. found that the diagnostic accuracy of QFR was lower in patients with high microcirculatory resistance, which is supported by our subgroup analysis. Due to its specific algorithm based on QCA and TIMI frame counting, coronary collateral circulation may reduce its accuracy. Therefore, corrective measures need to be developed to improve the diagnostic accuracy in patients with previous MI or high microcirculatory resistance. Furthermore, coronary calcification or thrombus may lead to angiographic




Fig. 4 - Summary receiver operating characteristic (sROC) curve for QFR.

haziness which undoubtedly reduces the diagnostic accuracy of QFR. A hybrid QFR-FFR approach may be a way to overcome these limitations. Yazaki, et al.^[6] found that FFR should be performed in stenosis with QFR 0.75-0.85. This hybrid approach may allow clinicians to get the best of both worlds by ensuring diagnostic accuracy while reducing cost and side effects.

It should be noted that our conclusion should be seen in the context of its limitation. First, the sample size is relatively small. Second, although there was no apparent heterogeneity in statistics, the heterogeneity in clinical and methodology was inevitable.

CONCLUSION

QFR is a simple, useful, and noninvasive modality for the diagnosis of functional significance of intermediate coronary artery stenosis.

No financial support.

No conflict of interest.

Authors' roles & responsibilities

ZX	Designed the study and provided methodological expertise; final approval of the version to be published
JP	Drafted the manuscript; final approval of the version to be published
JH	Drafted the manuscript; final approval of the version to be published
XH	Drafted the tables and figures; final approval of the version to be published
SG	Designed the study and provided methodological expertise; final approval of the version to be published

REFERENCES

1. van Nunen LX, Zimmermann FM, Tonino PA, Barbato E, Baumbach A, Engström T, et al. Fractional flow reserve versus angiography for guidance of PCI in patients with multivessel coronary artery disease (FAME): 5-year follow-up of a randomised controlled trial. *Lancet*. 2015 Nov 7;386(10006):1853-60. doi: 10.1016/S0140-6736(15)00057-4.
2. De Bruyne B, Pijls NH, Kalesan B, Barbato E, Tonino PA, Piroth Z, et al. Fractional flow reserve-guided PCI versus medical therapy in stable coronary disease. *N Engl J Med*. 2012 Sep 13;367(11):991-1001. doi:

- 10.1056/NEJMoa1205361. Epub 2012 Aug 27. Erratum in: N Engl J Med. 2012 Nov;367(18):1768. Möbius-Winckler, Sven [corrected to Möbius-Winkler, Sven].
3. Dattilo PB, Prasad A, Honeycutt E, Wang TY, Messenger JC. Contemporary patterns of fractional flow reserve and intravascular ultrasound use among patients undergoing percutaneous coronary intervention in the United States: insights from the National Cardiovascular Data Registry. *J Am Coll Cardiol*. 2012 Dec 4;60(22):2337-9. doi: 10.1016/j.jacc.2012.08.990.
 4. Tu S, Westra J, Yang J, von Birgelen C, Ferrara A, Pellicano M, et al. Diagnostic Accuracy of Fast Computational Approaches to Derive Fractional Flow Reserve From Diagnostic Coronary Angiography: The International Multicenter FAVOR Pilot Study. *JACC Cardiovasc Interv*. 2016 Oct 10;9(19):2024-35. doi: 10.1016/j.jcin.2016.07.013.
 5. Xu B, Tu S, Qiao S, Qu X, Chen Y, Yang J, et al. Diagnostic Accuracy of Angiography-Based Quantitative Flow Ratio Measurements for Online Assessment of Coronary Stenosis. *J Am Coll Cardiol*. 2017 Dec 26;70(25):3077-87. doi:10.1016/j.jacc.2017.10.035.
 6. Yazaki K, Otsuka M, Kataoka S, Kahata M, Kumagai A, Inoue K, et al. Applicability of 3-Dimensional Quantitative Coronary Angiography-Derived Computed Fractional Flow Reserve for Intermediate Coronary Stenosis. *Circ J*. 2017 Jun 23;81(7):988-992. doi: 10.1253/circj.CJ-16-1261.
 7. Liberati A, Altman DG, Tetzlaff J, Mulrow C, Gøtzsche PC, Ioannidis JP, et al. The PRISMA statement for reporting systematic reviews and meta-analyses of studies that evaluate healthcare interventions: explanation and elaboration. *BMJ*. 2009 Jul 21;339:b2700. doi:10.1136/bmj.b2700.
 8. Whiting P, Rutjes AW, Reitsma JB, Bossuyt PM, Kleijnen J. The development of QUADAS: a tool for the quality assessment of studies of diagnostic accuracy included in systematic reviews. *BMC Med Res Methodol*. 2003 Nov 10;3:25.
 9. Mejia-Renteria H, Lauri F, Lee JM, der Hoeven N, de Waard G, de Hoyps A, et al. TCT-70 Influence of microcirculatory resistance on the assessment of coronary stenosis severity with Quantitative Flow Ratio (QFR): Results of an international multicentre study. *J Am Coll Cardiol*. 2017;70(18 Suppl B): B30-1. doi: 10.1016/j.jacc.2017.09.121
 10. van Rosendaal AR, Koning G, Dimitriu-Leen AC, Smit JM, Montero-Cabezas JM, van der Kley F, et al. Accuracy and reproducibility of fast fractional flow reserve computation from invasive coronary angiography. *Int J Cardiovasc Imaging*. 2017 Sep;33(9):1305-1312. doi:10.1007/s10554-017-1190-3.
 11. Kameyama T, Kubo T, Emori H, Ino Y, Matsuo Y, Yamano T, et al. TCT-542 Usefulness of QFR measurement for non-culprit lesion of ACS patients. *J Am Coll Cardiol* 2016;68(18 Suppl B):B219. doi: 10.1016/j.jacc.2016.09.679
 12. Spitaleri G, Tebaldi M, Biscaglia S, Westra J, Brugaletta S, Enriquez A, et al. Quantitative Flow Ratio Identifies Nonculprit Coronary Lesions Requiring Revascularization in Patients With ST-Segment-Elevation Myocardial Infarction and Multivessel Disease. *Circ Cardiovasc Interv*. 2018 Feb;11(2):e006023. doi:10.1161/CIRCINTERVENTIONS.117.006023.
 13. Emori H, Kubo T, Kameyama T, Ino Y, Matsuo Y, Kitabata H, et al. Diagnostic Accuracy of Quantitative Flow Ratio for Assessing Myocardial Ischemia in Prior Myocardial Infarction. *Circ J*. 2018 Feb 23;82(3):807-814. doi:10.1253/circj.CJ-17-0949.
 14. Kobayashi Y, Lee JM, Fearon W, Lee J, Zimmermann F, Jung JH, et al. TCT-522 The Independence of Coronary Microvascular Dysfunction from Epicardial Disease Severity: Three-Vessel Invasive Coronary Physiologic Study. *J Am Coll Cardiol* 2016 (18 Suppl B):B210-211. doi 10.1016/j.jacc.2016.09.659
 15. Emori H, Kubo T, Kameyama T, Ino Y, Matsuo Y, Tanaka A, et al. TCT-529 The correlation between Quantitative Flow Ratio (QFR) and Fractional Flow Reserve (FFR). *J Am Coll Cardiol*. 2016;68(18 Suppl B):B213-4. doi: 10.1016/j.jacc.2016.09.666
 16. Tonino PA, Fearon WF, De Bruyne B, Oldroyd KG, Leesar MA, Ver Lee PN, et al. Angiographic versus functional severity of coronary artery stenoses in the FAME study fractional flow reserve versus angiography in multivessel evaluation. *J Am Coll Cardiol*. 2010 Jun 22;55(25):2816-21. doi: 10.1016/j.jacc.2009.11.096.
 17. Pijls NH, van Schaardenburgh P, Manoharan G, Boersma E, Bech JW, van't Veer M, et al. Percutaneous coronary intervention of functionally nonsignificant stenosis: 5-year follow-up of the DEFER Study. *J Am Coll Cardiol*. 2007 May 29;49(21):2105-11.
 18. Authors/Task Force members, Windecker S, Kolh P, Alfonso F, Collet JP, Cremer J, et al. 2014 ESC/EACTS Guidelines on myocardial revascularization: The Task Force on Myocardial Revascularization of the European Society of Cardiology (ESC) and the European Association for Cardio-Thoracic Surgery (EACTS) Developed with the special contribution of the European Association of Percutaneous Cardiovascular Interventions (EAPCI). *Eur Heart J*. 2014 Oct 1;35(37):2541-619. doi:10.1093/eurheartj/ehu278.

