



Iberoamerican Journal of Medicine

ISSN: 2695-5075

ISSN-L: 2695-5075

iberoamericanjm@gmail.com

Hospital San Pedro

España

Lo, Chi Ngai; Yeh, Ting-Ting; Tan, Chek Tien; Tsang, Eric W.
A reliability and validity study on upper limb range of motion
measurement using mobile sensor compared with goniometers
Iberoamerican Journal of Medicine, vol. 3, núm. 3, 2021, pp. 227-233
Hospital San Pedro
España

DOI: <https://doi.org/10.5281/zenodo.4904608>

Disponible en: <https://www.redalyc.org/articulo.oa?id=692072543008>

- Cómo citar el artículo
- Número completo
- Más información del artículo
- Página de la revista en redalyc.org

redalyc.org

Sistema de Información Científica Redalyc

Red de Revistas Científicas de América Latina y el Caribe, España y Portugal
Proyecto académico sin fines de lucro, desarrollado bajo la iniciativa de acceso
abierto



**IBEROAMERICAN
JOURNAL OF
MEDICINE**

iberoamericanjm

Journal homepage: www.iberoamericanjm.tk

Original article

A reliability and validity study on upper limb range of motion measurement using mobile sensor compared with goniometers

Chi Ngai Lo ^{a,*}, Ting-Ting Yeh ^a, Chek Tien Tan ^b, Eric W. Tsang ^c

^a Cluster of Health and Social Sciences, Singapore Institute of Technology, Singapore

^b Cluster of Infocomm Technology, Singapore Institute of Technology, Singapore

^c The Laboratory of Neuroscience for Education, Faculty of Education, The University of Hong Kong, Hong Kong

ARTICLE INFO

Article history:

Received 21 April 2021

Received in revised form

11 May 2021

Accepted 04 June 2021

Keywords:

Range of motion

Upper limb

Accelerometer

Goniometer

ABSTRACT

Introduction: The use of tele-rehabilitation devices to aid physiotherapy has gained popularity in recent years. In particular, measuring limb range of motion with a wearable mobile sensor can facilitate rehabilitation therapies by providing more efficient progress monitoring and reducing *clinicians' workload*. This study aimed to examine the test-retest reliability and validity of using a wearable mobile sensor to measure upper limb range of motion (ROM).

Materials and methods: Participants were recruited by convenient sampling. They were instructed to perform four kinds of upper limb movements including shoulder flexion, abduction, external rotation and elbow flexion, from which the ROM was measured by Mobile sensors REBEE (XCLR8 Technologies) and a standard goniometer (Model 12-1000) in each movement. Each kind of movements and the two ROM measurements were performed twice for the evaluation of test-retest reliability using Intraclass Correlation Coefficients (ICC). Pearson's correlations were computed between the ROM measured by the mobile sensors and the goniometer in each movement to assess construct validities of the mobile sensors. The agreement (95% mean difference) between the two sets of measurement was illustrated by Bland-Altman plots.

Results: Thirty-four asymptomatic young Asian adults (15 males) participated in this study ($M_{age} \pm SD$, 24.2 ± 3.82 years). The ICC for the ROM measured by the sensors were between 0.94 and 0.99, $p < 0.01$ and for the goniometer measurements were between 0.95 and 0.98, $p < 0.01$ in the four movements, indicating excellent reliability in both measurement methods. The *Pearson's correlation* between the sensor's and goniometric ROM measurements in four kinds of movements ranged from $r = 0.96$ to 0.99 , $p < 0.01$, indicating a very strong construct validity for using the mobile sensors to measure upper limb ROM. The mean difference between the two measurements ranged from 0.13 degrees to 7.6 degrees.

Conclusions: The results demonstrated that measuring upper limb ROM using the mobile sensors by non-healthcare trained students were as reliable and valid as using the standard universal goniometers by healthcare trained students. The findings implied that the wearable mobile sensors possibly allow non-healthcare trained carers to measure clinically useful, reliable and valid upper limb ROM from patients, which may facilitate symptom monitoring and improve the efficiency of the rehabilitation process.

© 2021 The Authors. Published by Iberoamerican Journal of Medicine. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>).

* Corresponding author.

E-mail address: lochingai@gmail.com

ISSN: 2695-5075 / © 2021 The Authors. Published by Iberoamerican Journal of Medicine. This is an open access article under the CC BY license

(<http://creativecommons.org/licenses/by/4.0/>).

<http://doi.org/10.5281/zenodo.4904608>

Estudio de confiabilidad y validez sobre la medición del rango de movimiento de las extremidades superiores utilizando un sensor móvil en comparación con goniómetros

INFO. ARTÍCULO

Historia del artículo:

Recibido 21 Abril 2021

Recibido en forma revisada

11 Mayo 2021

Aceptado 04 Junio 2021

Palabras clave:

Rango de movimiento

Miembro superior

Acelerómetro

Goniómetro

RESUMEN

Introducción: El uso de dispositivos de tele-rehabilitación para ayudar a la fisioterapia ha ganado popularidad en los últimos años. En particular, medir el rango de movimiento de las extremidades con un sensor móvil portátil puede facilitar las terapias de rehabilitación al proporcionar una supervisión del progreso más eficiente y reducir la carga de trabajo de los médicos. Este estudio tuvo como objetivo examinar la confiabilidad y la validez test-retest del uso de un sensor móvil portátil para medir el rango de movimiento de las extremidades superiores (ROM).

Materiales y métodos: Los participantes fueron reclutados mediante muestreo conveniente. Se les indicó que realizaran cuatro tipos de movimientos de las extremidades superiores, incluida la flexión del hombro, la abducción, la rotación externa y la flexión del codo, a partir de los cuales se midió el ROM con los sensores móviles REBEE (XCLR8 Technologies) y un goniómetro estándar (Modelo 12-1000) en cada movimiento. . Cada tipo de movimientos y las dos mediciones de ROM se realizaron dos veces para la evaluación de la confiabilidad test-retest utilizando coeficientes de correlación intraclase (ICC). Se calcularon las correlaciones de Pearson entre el ROM medido por los sensores móviles y el goniómetro en cada movimiento para evaluar la validez de constructo de los sensores móviles. La concordancia (diferencia media del 95%) entre los dos conjuntos de medidas se ilustró mediante gráficos de Bland-Altman.

Resultados: Treinta y cuatro adultos jóvenes asiáticos asintomáticos (15 hombres) participaron en este estudio (Mage $\pm$ SD, 24,2 $\pm$ 3,82 años). El ICC para el ROM medido por los sensores estuvo entre 0.94 y 0.99, $p < 0.01$ y para las mediciones del goniómetro estuvo entre 0.95 y 0.98, $p < 0.01$ en los cuatro movimientos, lo que indica una excelente confiabilidad en ambos métodos de medición. La correlación de Pearson entre el sensor y las mediciones de ROM goniométrica en cuatro tipos de movimientos varió de $r = 0,96$ a $0,99$, $p < 0,01$, lo que indica una validez de constructo muy fuerte para usar los sensores móviles para medir el ROM de las extremidades superiores. La diferencia media entre las dos medidas osciló entre 0,13 grados y 7,6 grados.

Conclusiones: Los resultados demostraron que la medición del ROM de las extremidades superiores mediante sensores móviles por parte de estudiantes no capacitados en salud era tan confiable y válida como el uso de goniómetros universales estándar por estudiantes capacitados en salud. Los hallazgos implicaron que los sensores móviles portátiles posiblemente permitan a los cuidadores no capacitados en atención médica medir el ROM de las extremidades superiores clínicamente útil, confiable y válido de los pacientes, lo que puede facilitar el monitoreo de los síntomas y mejorar la eficiencia del proceso de rehabilitación.

© 2021 Los Autores. Publicado por Iberoamerican Journal of Medicine. Éste es un artículo en acceso abierto bajo licencia CC BY (<http://creativecommons.org/licenses/by/4.0/>).

HOW TO CITE THIS ARTICLE: Lo CN, Yeh TT, Tan CT, Tsang EW. A reliability and validity study on upper limb range of motion measurement using mobile sensor compared with goniometers. Iberoam J Med. 2021;3(3):227-233. doi:

10.5281/zenodo.4904608.

1. INTRODUCTION

Clinical measurement of joint range of motion (ROM) has been a fundamental and essential evaluation in various health speciality, predominately physiotherapy. Its history can date back to the 1920-1930s [1]. Up to now, goniometry measurement for ROM is an everyday practice for physiotherapists as well as other healthcare professionals. To advance the ROM measurement, a new approach of using accelerometers and gyroscopes to facilitate ROM measurement has been intensively investigated in the last decade [2, 3].

An accelerometer is a force sensor that senses linear acceleration along with single or multiple directions. The operation principle is based on a mechanical sensing element that includes a proof mass attached to a mechanical suspension system. Following Newton's Second Law of Motion, the inertial force produced by gravity or acceleration causes the proof mass to deflect. This acceleration of the proof mass is quantified electrically [4]. A gyroscope detects angular motion about one or more axes. It operates on the mechanism of sensing an induced Coriolis acceleration on the proof mass causing vibration along a direction orthogonal to the axis about which the rotation is applied [5].

In recent years, there is a noticeably growing trend in the

use of telerehabilitation. With the use of mobile sensors, patients' physical status can be monitored remotely in the real world and in real-time [6], which increases efficiency and reduces the cost of rehabilitation services. Consequently, a reduction in healthcare disparities can be achieved. As the tele-rehabilitation system becomes more popular, biomedical engineers develop more variety of wearable sensors to facilitate the monitoring of patients' condition. Therefore, it is necessary to test the reliability and validity of the mobile sensors to ensure the accuracy of the measurement.

This study was to test the test-retest reliability of the human upper limb joints ROM measurement using a set of wearable mobile sensors. Also, it was to find out the validity compare with goniometer measurement.

2. MATERIALS AND METHODS

2.1. PARTICIPANTS

The participants for this study were students and staff of the Singapore Institute of Technology (SIT). They were recruited by convenient sampling. The inclusion criteria were healthy males or females aged 18 years old or above, no physical impairments in the upper limbs. Individuals not fulfilling the inclusion criteria were excluded. Ethical approval was issued by the Institutional Review Board of

the SIT (Approval number: 2018008). The study was conducted in Aug 2018 in the practical skill laboratory at the SIT and written informed consent was granted from the participants before the data collection.

2.2. INSTRUMENTS

The mobile sensor used in this study was a wearable REBEE (XCLR8 Technologies) movement detection system, which consists of a 10-axis inertial measurement unit including an accelerometer, a gyroscope and a magnetometer. The specifications of the sensor are as follows: (1) range parameters: Acceleration: 16g/s²; Gyro: 2000 deg/s; Bandwidth 20 Hz. (2) Data format: Quaternion data. (3) Communication parameters: Rate: 9600; Retrieval rate: 20 Hz. Two twelve-inch plastic BASELINE goniometers (Model 12-1000) were used in this experiment to compare the measurements with the sensors.

2.3. ACTIONS AND MEASUREMENTS

Four kinds of shoulder and elbow movements were studied for the measurement of ROM. Table 1 presents the details of these movements studied. These movements covered all the axis of movements in the shoulder and elbow and are functional in daily activities. The testing protocol followed the procedure described in Reese et al., 2016 [7]. The study was conducted at the exercise laboratory at SIT. During the

Table 1. Testing position of the four actions

Actions of range of motion	Subjects' position	Subjects' action	Sensors position
Shoulder flexion	The subject was asked to lie flat facing up on an examination plinth.	The testing arm rose up actively in the sagittal plane with the thumb of an opened palm pointing forwards.	Upper arm: The level just above the lateral epicondyle of humerus.
Shoulder abduction	Same position as above.	The testing arm rose up actively in the coronal plane with the thumb of an opened palm pointing out.	
Shoulder external rotation (in 90 degree abduction)	The subject was asked to lie flat facing up on an examination plinth with the testing arm put in 90 degree shoulder abduction, 90 degree elbow flexion and forearm pronation.	In the testing position, the forearm rose up actively in the sagittal plane with the shoulder abduction, elbow flexion and forearm pronation still.	Wrist: The level on the radial styloid process.
Elbow flexion	The subject was tested in the supine position with the palm of the arm pointing up. A folded towel was placed under the upper arm just above the elbow joint.	The subject bent the forearm fully in this position.	

data collection, the participants came into the laboratory and registered for the study. The investigators explained to the participants the purpose of the study. Participants who met the inclusion criteria of the study were invited to give their informed consents. The data was collected from participants individually. After recording their demographic information, the participant was randomly assigned to either the left limb or the right limb ROM measurement groups.

The investigators took the participant to the measuring station. The participants were instructed to lie in the supine position on the examination plinth. The sensors were placed in two regions in the upper arm and wrist to measure the joints ROM (Table 1). The two regions were selected because they are the most distal parts of the upper and lower arms with prominent bony landmarks. The sensors were stabilized on the testing limbs with elastic Velcro straps. The setup of the sensors was conducted by the two investigators (Team A) *who have a background in engineering* but were novice in healthcare and ROM knowledge. These two investigators also recorded the degrees of ROM readings from the system. After putting on the sensors, the participants were instructed about the movements as stated in Table 1. The participants were told to perform three to five repetitions of each movement as warm-up before the actual data recording. The participants were also instructed to perform each movement as far as possible and move the testing limb back to the starting position after the investigators took the readings from the sensors and the goniometers. Each of the four movements was repeated twice.

The control ROM measurements with the goniometers were taken by another two investigators (Team B) who were third-year students of a BSc. (Hons) Physiotherapy programme. They were supervised by two instructors who each had over ten years of clinical experience. The two investigators in Team A recorded the readings of the sensors for the angles of each of the four limb movements in degrees while the other two investigators in Team B conducted the goniometric measurements. Both sensors and goniometric measurements were conducted concurrently for each of the four movements studied. Still, all four investigators in both teams recorded the reading separately and were blinded from each other's readings.

2.4. STATISTICS

All data were analyzed with IBM SPSS 23 for Windows (IBM Corp., Armonk, NY, USA). Intraclass correlation coefficients (ICCs) and 95% confidence intervals (CIs) were computed to determine the test-retest reliability of the measurements. Pearson correlation coefficient (r) was calculated from the data between the sensors and goniometers to establish the ICCs and construct validity. Bland-Altman plots were generated to illustrate the 95% mean difference between the two measurements.

3. RESULTS

Thirty-four asymptomatic young Asian adults participated in this study and 15 of them were males. The average age of the participants was 24.2 ± 3.82 (mean $\pm$ standard deviation) years old. The average height and weight were 165.9 ± 8.3 cm and 61.2 ± 13.6 kg, respectively. Half ($N=17$) of participants were tested with their left limb and the other half on the right limb. A total of 544 data points were collected from the sensors and goniometer (34 participants $\times$ 4 movements $\times$ 2 rep. $\times$ 2 measurements). Among these, seven data points were discarded due to technical errors in the measurements with either the sensors or the goniometers. These included 2 data points in elbow flexion, 2 data points in shoulder flexion, 2 data points in shoulder abduction and 1 data point in shoulder external rotation.

The test-retest reliability as indicated by ICC for the ROM measured by the sensors were between 0.94 and 0.99, $p < 0.01$; and for the goniometer measurements were between 0.95 and 0.98, $p < 0.01$ in the four movements (Table 2). The reliability was considered as excellent strength (> 0.90) in both measurements, according to Koo and Li [10]. *Pearson's correlation* between the sensor's ROM and goniometric data in four kinds of movements ranged from $r = 0.96$ to 0.99 , $p < 0.01$, which explained 92% to 98% of variances indicating the correlation of these two measurements was very strong [11]. The Bland-Altman plots (Figure 1) illustrate that 95% of the differences in ROM measured by the sensors compared with the goniometers were from less than -10 deg to slightly more than 20 deg.

Table 2. Reliability of the sensors and goniometer measurement and correlation							
Action	Sensors		Goniometer		Mean Difference [95% CI]	r value	R- squared
	Mean $\pm$ SD	ICC _{2,1} [95% CI]	Mean $\pm$ SD	ICC _{2,1} [95% CI]			
Shoulder flexion (n = 32)	165.8 $\pm$ 7.8	*0.94 [0.89-0.97]	165.9 $\pm$ 9.4	*0.95 [0.89-0.97]	-0.13 [-2.2-2.0]	**0.99	0.98
Shoulder Abduction (n = 33)	166.3 $\pm$ 11.5	*0.99 [0.98-0.99]	173.9 $\pm$ 12.7	*0.98 [0.97-0.99]	-7.6 [-10.2-4.9]	**0.99	0.98
Shoulder external rotation (in 90 degree abduction) (n = 32)	94.7 $\pm$ 12.7	*0.96 [0.92-0.98]	89.5 $\pm$ 10.9	*0.98 [0.97-0.99]	5.2 [2.7-7.7]	**0.96	0.92
Elbow flexion (n=32)	145.7 $\pm$ 8.9	*0.98 [0.95-0.99]	142.0 $\pm$ 8.3	*0.96 [0.92-0.98]	3.7 [1.1-6.4]	**0.98	0.96

The reference ROM values of shoulder flexion, abduction, external rotation (in 90 deg abduction) and elbow flexion are 170.4, 184.5, 98.6 and 147.3 respectively for healthy individuals in the age group 20-40 years old [8, 9].

ICC = Intraclass correlation coefficient. * ICC_{2,1} with $p < 0.01$. **r value with $p < 0.01$




Figure 1: Bland-Altman plots of results of two measurements in shoulder flexion (A), shoulder abduction (B), shoulder external rotation (C) and elbow flexion (D).

4. DISCUSSION

Movement measurements through accelerometers are not a novel idea. Several previous studies had been conducted to test for the accuracy of the measurement devices [12, 13]. In the current study, the specifications of the sensors and the measurement procedures were stated clearly, which allowed future studies to reproduce the procedures in further testing. With the advancement of ROM measurement technology, the findings of this study showed that investigators with no prior training in ROM measurement could use the mobile sensors and achieve highly reliable and valid ROM measurements compared to those of health professionals, which further indicated the practicability of this technology.

Similar to the previous studies of measuring joint angles using electronic sensors [2, 14], the mobile sensors used in this study demonstrated excellent test-retest reliability of upper limb ROM measurements. The measurements of upper limb ROM data between the sensors and goniometers in all four kinds of movements had a very strong correlation, which was especially noteworthy due to our effective blinding process in our measurement procedure. The mean difference between the two measurements and the Bland–Altman plots provide more specific ideas about the correlation. The ROM measurements of the sensors and the goniometer were concurrently but separately recorded by two teams of investigators who were not aware of the other team's readings. In addition, the team A investigators who operated the sensors were not trained in healthcare knowledge and therefore, they had no preconception of the sensor's ROM readings. It was previously proposed that the reliability of goniometry depends on the clinicians' experience [15, 16]. Therefore, our findings implied that novice investigators could operate the sensors properly and performed reliably with valid clinical ROM measurements. The mobile sensor system may allow *non-healthcare trained* family members to record clinically useful measurements for upper limb ROM for symptom monitoring when therapists were not present to tend for the patients.

The average values in all four kinds of ROM were close to the previously established normative ROM values of the shoulder joint from the same age group [8, 9], which provided support for the validity of our ROM data. The mean differences between the two ROM measurements for shoulder flexion and shoulder abduction were 0.13 degrees and 7.6 degrees respectively. The mean difference and 95% CI indicated that there was no statistically significant

difference. The ROM of the elbow flexion and shoulder external rotation movements recorded by the sensors were statistically significantly higher than that of the goniometer, as indicated by the mean difference 95% CI (Table 2). Traditionally, goniometry is a gold standard for measuring shoulder ROM [17] but not for elbow ROM. Instead, the gold standard for elbow ROM is radiographic measurement [18]. However, we found that the correlations of the upper limb ROM measurements between the sensor and goniometer in all four kinds of movements were over 0.95 [18], indicating that the measurement of elbow ROM with goniometers is a valid method.

This study did not include lower limbs measurement for a few reasons. First of all, the ROM measurements recorded by the standardized procedures during lower limb movements such as knee flexion or single-leg raises were not as functional. Some clinicians and researchers would prefer taking the ROM of the lower limb in more functional ways, such as standing and walking [2, 19]. Secondly, the sensors were developed for telerehabilitation, for which the patients are supposed to put on the sensors and perform the actions themselves, or with assistance from non-healthcare trained people. Without the supervision of clinicians or researchers, the lower limb ROM measurements using the standard procedures are certainly inaccurate due to many possible measurement pitfalls [19, 20]. The measurement of lower limb ROM with the sensors using the standard procedures would be difficult for non-healthcare trained investigators to perform. Accordingly, the current study focused on establishing the reliability and validity of ROM measurements of the upper limb using the mobile sensors. Convenient sampling is one of the limitations of this study. The current study recruited merely healthy and young participants; the generalizability of our results to different populations of participants would be limited.

5. CONCLUSIONS

In conclusion, the wearable mobile sensors REBEE were found to be a reliable and valid device for upper limb ROM measurement. This mobile sensor can be a comparable alternative to the goniometers to measure human upper limbs ROM in clinical practice and research.

6. CONFLICT OF INTEREST

The equipment and payment for three physiotherapy student investigators (less SGD 500) of this study were

provided by XCLR8 Technologies.

The authors did not receive any financial or non-financial support.

7. REFERENCES

1. Gajdosik RL, Bohannon RW. Clinical measurement of range of motion. Review of goniometry emphasizing reliability and validity. *Phys Ther.* 1987;67(12):1867-72. doi: 10.1093/ptj/67.12.1867.
2. Dejnabadi H, Jolles BM, Aminian K. A new approach to accurate measurement of uniaxial joint angles based on a combination of accelerometers and gyroscopes. *IEEE Trans Biomed Eng.* 2005;52(8):1478-84. doi: 10.1109/TBME.2005.851475.
3. Noorkõiv M, Rodgers H, Price CI. Accelerometer measurement of upper extremity movement after stroke: a systematic review of clinical studies. *J Neuroeng Rehabil.* 2014;11:144. doi: 10.1186/1743-0003-11-144.
4. Yang CC, Hsu YL. A review of accelerometry-based wearable motion detectors for physical activity monitoring. *Sensors (Basel).* 2010;10(8):7772-88. doi: 10.3390/s100807772.
5. Seshia AA, Howe RT, Montague S. An integrated microelectromechanical resonant output gyroscope. Technical Digest. MEMS 2002 IEEE International Conference. Fifteenth IEEE International Conference on Micro Electro Mechanical Systems (Cat. No.02CH37266). 2002;722-6. doi: 10.1109/MEMSYS.2002.984372.
6. Dobkin BH. Wearable motion sensors to continuously measure real-world physical activities. *Curr Opin Neurol.* 2013;26(6):602-8. doi: 10.1097/WCO.0000000000000026.
7. Reese N, Bandy W. *Joint Range of Motion and Muscle Length Testing*. 3rd ed. Saunders; 2016.
8. Soucie JM, Wang C, Forsyth A, Funk S, Denny M, Roach KE, Boone D; Hemophilia Treatment Center Network. Range of motion measurements: reference values and a database for comparison studies. *Haemophilia.* 2011;17(3):500-7. doi: 10.1111/j.1365-2516.2010.02399.x.
9. Barnes CJ, Van Steyn SJ, Fischer RA. The effects of age, sex, and shoulder dominance on range of motion of the shoulder. *J Shoulder Elbow Surg.* 2001;10(3):242-6. doi: 10.1067/mse.2001.115270.
10. Koo TK, Li MY. A Guideline of Selecting and Reporting Intraclass Correlation Coefficients for Reliability Research. *J Chiropr Med.* 2016;15(2):155-63. doi: 10.1016/j.jcm.2016.02.012.
11. Schober P, Boer C, Schwarte LA. Correlation Coefficients: Appropriate Use and Interpretation. *Anesth Analg.* 2018;126(5):1763-8. doi: 10.1213/ANE.0000000000002864.
12. McGinnis RS, Patel S, Silva I, Mahadevan N, DiCristofaro S, Jortberg E, et al. Skin mounted accelerometer system for measuring knee range of motion. *Annu Int Conf IEEE Eng Med Biol Soc.* 2016;2016:5298-302. doi: 10.1109/EMBC.2016.7591923.
13. Mourcou Q, Fleury A, Franco C, Klopčič F, Vuillerme N. Performance Evaluation of Smartphone Inertial Sensors Measurement for Range of Motion. *Sensors (Basel).* 2015 15;15(9):23168-87. doi: 10.3390/s150923168.
14. El-Gohary M, McNames J. Shoulder and elbow joint angle tracking with inertial sensors. *IEEE Trans Biomed Eng.* 2012;59(9):2635-41. doi: 10.1109/TBME.2012.2208750.
15. Lewis E, Fors L, Tharion WJ. Interrater and intrarater reliability of finger goniometric measurements. *Am J Occup Ther.* 2010;64(4):555-61. doi: 10.5014/ajot.2010.09028.
16. van de Pol RJ, van Trijffel E, Lucas C. Inter-rater reliability for measurement of passive physiological range of motion of upper extremity joints is better if instruments are used: a systematic review. *J Physiother.* 2010;56(1):7-17. doi: 10.1016/s1836-9553(10)70049-7.
17. Werner BC, Holzgrefe RE, Griffin JW, Lyons ML, Cosgrove CT, Hart JM, et al. Validation of an innovative method of shoulder range-of-motion measurement using a smartphone clinometer application. *J Shoulder Elbow Surg.* 2014;23(11):e275-82. doi: 10.1016/j.jse.2014.02.030.
18. Chapleau J, Canet F, Petit Y, Laflamme GY, Rouleau DM. Validity of goniometric elbow measurements: comparative study with a radiographic method. *Clin Orthop Relat Res.* 2011;469(11):3134-40. doi: 10.1007/s11999-011-1986-8.
19. Vohralik SL, Bowen AR, Burns J, Hiller CE, Nightingale EJ. Reliability and validity of a smartphone app to measure joint range. *Am J Phys Med Rehabil.* 2015;94(4):325-30. doi: 10.1097/PHM.0000000000000221.
20. Clarkson HM. *Musculoskeletal assessment: joint range of motion and manual muscle strength*. 2nd ed. Lippincott Williams & Wilkins; 2000.