



Anais da Academia Brasileira de Ciências

ISSN: 0001-3765

[aabc@abc.org.br](mailto:aabc@abc.org.br)

Academia Brasileira de Ciências

Brasil

Fernández-Morales, Flavio H.; Duarte, Julio E.; Samitier-Martí, Josep  
Bacterial handling under the influence of non-uniform electric fields: dielectrophoretic and  
electrohydrodynamic effects  
Anais da Academia Brasileira de Ciências, vol. 80, núm. 4, diciembre, 2008, pp. 627-638  
Academia Brasileira de Ciências  
Rio de Janeiro, Brasil

Available in: <http://www.redalyc.org/articulo.oa?id=32713465004>


- How to cite
- Complete issue
- More information about this article
- Journal's homepage in [redalyc.org](http://redalyc.org)

[redalyc.org](http://redalyc.org)

Scientific Information System

Network of Scientific Journals from Latin America, the Caribbean, Spain and Portugal

Non-profit academic project, developed under the open access initiative



Anais da Academia Brasileira de Ciências (2008) 80(4): 627-638  
(Annals of the Brazilian Academy of Sciences)  
ISSN 0001-3765  
www.scielo.br/aabc

## Bacterial handling under the influence of non-uniform electric fields: dielectrophoretic and electrohydrodynamic effects

FLAVIO H. FERNÁNDEZ-MORALES<sup>1</sup>, JULIO E. DUARTE<sup>1</sup> and JOSEP SAMITIER-MARTÍ<sup>2</sup>

<sup>1</sup>Grupo de Energía y Aplicación de Nuevas Tecnologías (GEANT), Seccional Duitama

Universidad Pedagógica y Tecnológica de Colombia, Carrera 18, Calle 22, Duitama, Boyacá, Colombia

<sup>2</sup>Departamento de Electrónica, Universidad de Barcelona, c./ Martí i Franqués 1, 08028, Barcelona, España

*Manuscript received on June 19, 2007; accepted for publication on August 18, 2008;  
presented by LUIZ DAVIDOVICH*

### ABSTRACT

This paper describes the modeling and experimental verification of a castellated microelectrode array intended to handle biocells, based on common dielectrophoresis. The proposed microsystem was developed employing platinum electrodes deposited by lift-off, silicon micromachining, and photoresin patterning techniques. Having fabricated the microdevice it was tested employing *Escherichia coli* as bioparticle model. Positive dielectrophoresis could be verified with the selected cells for frequencies above 100 kHz, and electrohydrodynamic effects were observed as the dominant phenomena when working at lower frequencies. As a result, negative dielectrophoresis could not be observed because its occurrence overlaps with electrohydrodynamic effects; i.e. the viscous drag force acting on the particles is greater than the dielectrophoretic force at frequencies where negative dielectrophoresis should occur. The experiments illustrate the convenience of this kind of microdevices to microhandling biological objects, opening the possibility for using these microarrays with other bioparticles. Additionally, liquid motion as a result of electrohydrodynamic effects must be taken into account when designing bioparticle micromanipulators, and could be used as a mechanism to clean the electrode surfaces, that is one of the most important problems related to this kind of devices.


**Key words:** dielectrophoresis, electrohydrodynamic effects, biochips, bacterial handling.

### INTRODUCTION

Handling of biological objects constitutes an important operation in many biotechnological processes such as biochemical assays, biocell detection and separation, cell fusion and gene manipulation (Hoettges et al. 2003, Figeys and Pinto 2000, Lee and Tai 1999). In conventional techniques, cells are often treated as a mass in the form of suspensions in a liquid medium. However, cells have dimensions in the order of microns and are usually fragile, and mechanical components commonly employed to perform this handling are too large and stiff (Sato et al. 1990).

Diverse non-contact techniques have been proposed using electric, magnetic, ultrasonic and other forces as the actuation mechanism to handle and characterize biological cells (Choi et al. 1999, Fuh et al. 1998a, Porras et al. 2004). Among these alternatives, electrostatic fields are the most suitable for miniaturization because they do not have mechanical moving parts, and only a few electrodes have to be fabricated (Sato and Shirley 1998).

Electric fields involve two phenomena. The first one is electrophoresis that uses direct current voltage to achieve migration of charged particles in a solution.



second one is dielectrophoresis, DEP, that is more suitable because dissociation is prevented using high-frequency signals.

DEP was a term first used by Pohl in 1951 in order to describe motion of electrical neutral matter caused by its interaction with non-uniform electric fields (Pohl 1951). This phenomenon has been employed in applications such as: artificial-particle and bioparticle characterization (Burt et al. 1990), devices for cell fusion (Lee et al. 1995), particle separation based on its differential polarizability (Gascoyne et al. 1997), particle collection and patterning (Velev and Kaler 1999, Rosentel and Voldman 2005), cellular culture over a field-protected electrode surface (Fuhr and Wagner 1994) and particle handling in terms of linear motion and positioning (Wang et al. 1997), levitation (Kaler and Pohl 1983), and trapping in electric-field cages (Fuhr et al. 1998b).

In the past 25 years lots of engineering work has been invested to develop bioparticle microhandling microtools, that implies the fabrication of devices whose size roughly matches that of the bioparticle itself, normally between 1 to 100  $\mu\text{m}$  or less if virus or macromolecules must be handled. Nowadays these sizes and resolution levels may only be achieved using microsystems technology that suits very well the requirements of miniaturized microstructures aimed to handle biological objects in order to obtain more confident, faster and cheaper biochemical assays (Reimer et al. 1995). Besides diverse materials and technological approaches have been used to fabricate these microtools, silicon substrates with electrodes made in noble metals are usually employed to manufacture such devices (Pethig et al. 1998, Schnelle et al. 1996).

This paper describes the modeling and experimental verification of a castellated microelectrode array intended to handle biocells, based on common dielectrophoresis. The proposed microsystem was developed employing platinum electrodes deposited by lift-off, silicon micromachining, and photoresin patterning techniques. Having fabricated the microdevice it was tested employing *Escherichia coli* as bioparticle model.

Positive DEP could be verified with the selected

DEP could not be experimentally verified because electrohydrodynamic effects, usually unconsidered, were observed as the dominant phenomena when working at lower frequencies. It shows that the liquid motion as a result of electrohydrodynamic effects must be taken into account when designing bioparticle micromanipulators, and could be used as mechanism to clean the electrode surfaces, that is one of the most important problems related to this kind of devices. The most relevant results as well as the basic phenomena involved in microhandling of bacteria by means of non-uniform electric fields are reported underneath.

## MATERIALS AND METHODS

### THE DIELECTROPHORETIC FORCE

Dielectrophoresis, DEP, is defined as the lateral motion imparted on uncharged particles as a result of polarization induced by non-uniform electric fields. This motion may be toward or away from the electric field maximum, depending on the electrical properties of the particles and suspending medium, respectively. Such a definition stands for the so-called conventional or common dielectrophoretic (c-DEP) phenomena.

For a non-ideal insulating spherical particle of radius  $r$  and relative complex dielectric permittivity  $\varepsilon_p^*$  suspended in a medium of relative complex dielectric permittivity  $\varepsilon_m^*$  and interacting with an electric field of strength  $\mathbf{E}$ , the time-averaged dielectrophoretic force is given by (Wang et al. 1993):


$$\mathbf{F}_{\text{DEP}} = 2\pi \varepsilon_0 \varepsilon_m r^3 \text{Re}[\mathbf{F}_{\text{CM}}] \nabla |\mathbf{E}|^2 \quad (1)$$

where  $\varepsilon_0 = 8.854 \times 10^{-12}$  (Farad  $\text{m}^{-1}$ ) is the free-space permittivity,  $\nabla$  is the gradient operator and  $\text{Re}[\mathbf{F}_{\text{CM}}]$  denotes the real part of the Clausius-Mossotti factor defined by:

$$\mathbf{F}_{\text{CM}} = \frac{(\varepsilon_p^* - \varepsilon_m^*)}{(\varepsilon_p^* + 2\varepsilon_m^*)}. \quad (2)$$

The relative complex dielectric permittivity is given by:

$$\varepsilon^* = \varepsilon_0^\varepsilon - j^{\sigma/\omega} \quad (3)$$



## BACTERIAL HANDLING BY NON-UNIFORM ELECTRIC FIELDS

DEP can be divided into two phenomena, as depicted in Figure 1. The first one is known as positive dielectrophoresis (p-DEP). It is characterized by the particle motion towards the strongest, or maximum, region of the electric field as a consequence of a positive value of the  $\text{Re}[F_{\text{CM}}]$ . p-DEP occurs for particles that are more polarizable than the medium. On the other hand, if the particle polarizability is low enough for the  $\text{Re}[F_{\text{CM}}]$  to become negative, i.e. when the particle effective permittivity and/or conductivity are smaller than those of the medium, the  $\mathbf{F}_{\text{DEP}}$  will tend to move the particle towards the weakest, or minimum, region of the electric field. This phenomenon is called negative dielectrophoresis (n-DEP) (Stephens et al. 1996).




Fig. 1 – Two different particles in a non-uniform electric field. The particle on the left is more polarizable than the surrounding medium and is attracted towards the strong field at the pin electrode (p-DEP), whilst the particle of low polarizability on the right is directed away from the strong field region (n-DEP).

From equation (1) it can be seen that  $\mathbf{F}_{\text{DEP}}$  depends on the particle size, as well as on the magnitude of the electric field which is related to the electric potential applied. The gradient operator stands for the spatial non-uniformity of the electric field applied, which is a geometrical factor depending on the electrode layout (Feeley and Pohl 1981). Furthermore,  $\mathbf{F}_{\text{DEP}}$  can actuate on both direct and alternating current fields as a consequence of

of this dielectrophoretic force depends on the polarizability of the induced dipole moment, which in turn is determined by the conductivities and permittivities of the particle and its suspending medium.

Thus, one of the most important parameters that governs DEP is the polarizability function, the so-called Clausius-Mossotti factor, because it establishes the direction of the dielectrophoretic force, i.e.  $F_{\text{CM}}$  determines the occurrence of p-DEP or n-DEP depending on whether  $\text{Re}[F_{\text{CM}}]$  is positive or negative, respectively.

Equation (2) corresponds to the homogeneous sphere model that can be used with polystyrene beads (Arnold et al. 1987). However, when the behavior of cells such as viruses, bacteria or protoplasts is to be explained, the single-shell particle model is a better choice. Living cells are usually modeled as a homogeneous, permittivities, spherical particle of radius  $r$  having complex permittivity  $\epsilon_{\text{int}}^*$ , surrounded by a very thin shell corresponding to a membrane of thickness  $d$  having complex permittivity  $\epsilon_{\text{mem}}^*$ , which defines the effective complex permittivity as (Gascoyne et al. 1995):


$$\epsilon_p^* = \epsilon_{\text{mem}}^* \frac{\left(1 + \frac{d}{r}\right)^3 + 2 \left[ \frac{(\epsilon_{\text{int}}^* - \epsilon_{\text{mem}}^*)}{(\epsilon_{\text{int}}^* + 2\epsilon_{\text{mem}}^*)} \right]}{\left(1 + \frac{d}{r}\right)^3 - \left[ \frac{(\epsilon_{\text{int}}^* - \epsilon_{\text{mem}}^*)}{(\epsilon_{\text{int}}^* + 2\epsilon_{\text{mem}}^*)} \right]}$$

Replacing equation (4) in the expression for the Clausius-Mossotti factor, equation (2), it is possible to obtain further insight into the behavior of bioparticles under the influence of non-uniform electric fields.

## THE ELECTRIC FIELD MODELING

Equation (1) shows that  $\mathbf{F}_{\text{DEP}}$  depends on three factors: the particle size, the electric properties of particle and medium, and the geometrical inhomogeneities of the electric field ( $\nabla E_{\text{RMS}}^2$ ). Furthermore, the last one hinges on the physical layout of the selected electrodes, while the other ones are independent of such electrode considerations. In view of this, it is necessary to calculate the electric field over the electrode surface.

Calculations presented below correspond to a periodic interdigitated castellated electrode layout densely



uses the Finite Element Method (FEM) to solve physical problems (Kohnke 1995).




Fig. 2 – Classical interdigitated castellated microelectrode pattern, solved by a FEM model with a castellation size and medium height of  $50\mu\text{m}$  and an electrode thickness of  $1\mu\text{m}$ .

Having applied a difference potential of  $\pm 5\text{ V}$  and considering the interelectrode surface gap as ground, the electric field profile was assessed giving as a result the field-modulus distribution represented in Figure 3. For the sake of simplicity, only a plane located at  $2\mu\text{m}$  over the electrode surface is analyzed. It can be seen that points located near to the electrode tips will experience the highest electric field strength, whilst this strength diminishes as the observation point moves towards the central part of the interelectrode gap, towards the bays or onto the metallic surface of the electrodes. As a result, particles more polarizable than the medium will be collected at the electrode tips. On the other hand, particles less polarizable than the suspending medium will be repelled from the field maxima, clustering at the bay zones or on the electrodes. If  $\mathbf{E}$  is represented at higher planes its strength will decay very fast, almost exponentially.

Additionally, a cross-section view of the geometrical inhomogeneity factor ( $\nabla E_{\text{RMS}}^2$ , which is proportional to  $\mathbf{F}_{\text{DEP}}$ ) is plotted in Figure 4. For the sake of

regions are located around the corners of the electrodes and, as a result of this, the highest value of  $\mathbf{F}_{\text{DEP}}$  will be located at these points.

Particles lying in the force field shown in Figure 4 will present p-DEP and will be attracted to and agglomerated at the highest strength region of the electric field if the polarizability function is positive. If particles were oppositely polarized, they would be repelled towards the force field minima as a result of n-DEP.

To sum up, the computer analysis predicts that p-DEP will result in the agglomeration of particles at electrode tips, while n-DEP will cause particles to accumulate on electrode bays and on the electrode surface, as well as in the interelectrode gap. Of course, the final particle position depends upon the polarizability function value.


One of the most amazing possibilities envisaged for the castellated microelectrodes is that multiple-particle mixtures (two or even more particle types) could be physically separated into different regions, on the basis of their electrical properties. Besides this, an additional force can be used to drag away the desired fraction of particles, collecting them for further assays. However, one must be aware that such a separation involves a careful choice of the experimental conditions, especially the suspending medium conductivity.

#### THE PROPOSED MICRODEVICE

The proposed design is a whole microsystem including electrical, optical and fluidic interfaces, as depicted in Figure 5.

The technological process was carried out at the Microelectronics National Center (CNM) in Barcelona, Spain, and can be roughly divided into three stages. The first one is the microelectrode patterning in which platinum electrodes are defined by lift-off onto a silicon wafer of  $300\mu\text{m}$  thickness. After that, the wafer is drilled by bulk silicon micromachining in order to shape the inlet and outlet holes. The wafer-level processing ends up with the photolithographic structuring of an UV-curable polymer to cast the microchamber walls.

As shown in Figure 6, the resulting microelectrode



# BACTERIAL HANDLING BY NON-UNIFORM ELECTRIC FIELDS





Fig. 3 – Pseudo-topographic representation of the electric field modulus ( $V \mu m^{-1}$ ) in a plane located at  $2\mu m$  over the electrode surface.







Fig. 5 – Cross-section view of the proposed microdevice. Three main components may be identified: the substrate where electrodes are grown, the electrodes itself responsible for the electric field profiles, and the silicon or photoresin walls limiting the microchamber working area. A couple of holes are drilled to bring the suspending medium onto the electrode surface. A cover slide is placed on the top of the structure to close the cavity.

with typical sizes of 50, 70 and 90  $\mu\text{m}$  in both electrode length and separation. In order to gain flexibility, electrodes of different size have their own pads for external connection.





Fig. 6 – Platinum microstructures containing classical and shifted in-

#### EXPERIMENTAL SET-UP

The instrumentation required to perform cell microhandling can be separated into three parts. The first one is the power system that includes a signal generator and an oscilloscope. In this case, the power signal was supplied by a Hewlett Packard 33120 A function generator with amplitude variable between 0 and 10 V and a frequency range from 0 to 20 MHz. The electrical signal calibration was allowed by means of a Tektronix TSD 220, 100 MHz, two channel digital real-time oscilloscope.

The second part is the particle microhandler itself, which contains the microelectrodes under study. The third part is the optical system that is able to observing and recording the electrokinetic particle behavior. The experimental work was performed employing a stereomicroscope Karl Zeiss SV 11 with a Sony CCD video color camera SSC-370 P, as well as a videotape recorder attached to it. Moreover, there was a color video monitor, Sony KX-1410 QM, connected in parallel to supervise the experiments. Furthermore, taking advantage of a PC provided with a professional quality PCI



## BACTERIAL HANDLING BY NON-UNIFORM ELECTRIC FIELDS

of them. This system is complemented by a cold-light source, which allows light filtering and polarization. Additionally, the liquid medium conductivity was measured by means of a Corning® 441 conductivity meter. The whole experimental set-up is placed on an anti-vibration table to reduce the influence of possible external movements when observing the particle electrokinetic behavior (Fernández et al. 2002).

### RESULTS

Having fabricated the microdevice it was ready to demonstrate the major goal of this research: bioparticle microhandling by means of microsystems based on dielectrophoresis. To do this *Escherichia coli* were used as bioparticle models. They correspond to gram-negative bacterial cells which have been widely studied and their electrical properties are well known (Markx et al. 1994, Brown 1996, Cheng et al. 1998). Results described underneath were obtained employing interdigitated castellated microelectrodes of  $70\mu\text{m}$  in size. Such structures are well-suited to study c-DEP phenomena because they are only driven by two signals inverted in phase.

Gram-negative bacteria (*E. coli*) were kindly provided by the Microbiology Department of the Barcelona University. All measurements were performed in very low conductivity solutions to reduce the risk of heating-related problems. Bioparticles were resuspended in de-ionized water and the final suspension was adjusted to a concentration of  $2 \times 10^8$  cells  $\text{mL}^{-1}$ . The solution conductivity was adjusted between 10 and  $20\mu\text{S cm}^{-1}$ . The shape of these cells is usually tubular with the long dimension varying from submicrometer to about  $4\mu\text{m}$ .

Having dropped the bacterial suspension into the micropool, microelectrodes of  $70\mu\text{m}$  were energized with a sinusoidal signal of 10 V in amplitude and 1 MHz. After a few minutes, particles were concentrated forming clusters located at the electrode edges of both classical and shifted interdigitated castellated arrays, i.e. particles were driven towards and agglomerated at regions corresponding to the electric field maxima, as predicted by simulations of the electric field. The aforementioned

ters of trapped bioparticles (bright points) onto the interelectrode gap and bays are free of particles. *E. coli* preferentially agglomerate at the electrode edges as well as at the back-side of the electrode bays.





Fig. 7 – Sequence of bacterial movement over the interdigitated microstructure. The applied voltage was 10 V. Frames were captured at: (a) 100 kHz, (b) 20 kHz, (c) 10 kHz, and (d) 1 kHz. Bacteria were diluted in de-ionized water ( $\sigma_m = 10\mu\text{S cm}^{-1}$ ), Magnification  $\times 200$ .

When the frequency of the applied signal was increased till approximately 100 kHz, maintaining constant the voltage amplitude, bioparticle behavior (direction of motion towards and accumulation at the electrode edges) remained unaltered. The signal frequency was raised till the highest value allowed by our generation system (20 MHz) but n-DEP could not be verified. In other words, the cell aggregation previously described was observed from 100 kHz to 20 MHz, indicating that for our experimental conditions only p-DEP could be observed. When the applied signal was switched off, bioparticles were observed to redistribute randomly in the micropool as they were prior to applying the electrical signal.

Once bacteria had been collected at the electrode edges under the influence of p-DEP, the applied frequency was reduced below 100 kHz till a value of 1 kHz. As the frequency was decreased, particles separated from the electrode edges and moved towards the center, accumulating to form diamond-shaped clusters.





ior, i.e. particle collection at the center of the electrode surfaces when working at relatively low frequencies in dielectrophoretic-based microdevices, has also been observed by other researchers when working in handling, characterization and separation of microparticles (Green and Morgan 1998, Green et al. 2000a, b).

## DISCUSSION

### POSITIVE DIELECTROPHORESIS

To explain the bioparticle behavior described in Figure 7a, the single-shell particle model was fitted with values corresponding to our experimental choice, i.e. *E. coli* suspended in de-ionized water of  $\sigma_m = 1 \times 10^{-3} \text{ S m}^{-1}$  and  $\epsilon_m = 80$ . A plot of the real part of the polarizability function is shown in Figure 8, in which the bioparticle values were taken from the literature (Nishioka et al. 1997, Hughes and Morgan 1999).




Fig. 8 – Magnitude of  $\text{Re}[F_{\text{CM}}]$  versus frequency for an *E. coli* of  $\epsilon_{\text{mem}} = 3.7$ ,  $\epsilon_{\text{int}} = 50$ ,  $\sigma_{\text{mem}} = 0 \mu\text{S m}^{-1}$  and  $\sigma_{\text{int}} = 0.5 \text{ S m}^{-1}$ , submersed in de-ionized water of  $\sigma_{\text{mem}} = 1 \times 10^{-3} \text{ S m}^{-1}$ . Moreover,  $d = 10 \text{ nm}$  and  $r = 2 \mu\text{m}$ .

A careful study of Figure 8 reveals that  $\text{Re}[F_{\text{CM}}]$  has three well-defined behavioral frequency regions. The first one ranges from 1 to about 30 kHz, in which n-DEP will occur and is limited by a low-frequency dispersion which gives rise to the first crossover point. The second region ranges from 30 kHz to approximately 100 MHz at which a second crossover point can be seen. A third region, within which n-DEP will occur, ranges

30 kHz to 100 MHz, which is consistent with our experimental observations. On the other hand, Figure 8 shows that to verify n-DEP frequencies lower than 30 kHz or higher than 100 MHz must be applied. However, while the latter range was not available, the former triggered EHD phenomena hindering bioparticles from behaving in the n-DEP regime.


In view of the experimental results one can say that the occurrence of p-DEP in *Escherichia coli* was successfully experimentally verified. Furthermore, cell behavior is coherent with that described by the particle model corresponding to a single-shell equivalent sphere for frequencies higher than 100 kHz.

### ELECTROHYDRODYNAMIC EFFECTS

Since the point of view of ‘classical’ dielectrophoretic theory, *E. coli* should remain trapped at the electrode edges for frequencies ranging from 30 kHz to 100 MHz as a result of p-DEP, because the real part of the polarizability function of bacteria is positive in this range. For frequencies below 30 kHz, they would be directed toward the electrode bays and interelectrode gaps but such a prediction of the single-shell model could not be verified. On the contrary, it is evident that particles in the interelectrode regions literally jump onto the electrode surface and then move, together with those cells originally on the electrode surface, towards the central region, as shown in Figure 7.

Such a behavior was tentatively attributed to the combination of an electrophoretic effect in aiding cells to cross the electrode boundaries, coupled with a negative dielectrophoretic force in directing them to the center of the electrode surface (Pethig et al. 1992). However, the impressive review of Ramos and co-workers worked out an order-of-magnitude estimation of forces acting in c-DEP microelectrodes, concluding that for the low-frequency range the fluid motion moves particles away from the electrode edge and into well defined regions on top of the electrodes (Ramos et al. 1998).

In fact, the aforementioned review aimed us to perform the numerical analysis of electrohydrodynamic (EHD) effects, i.e. fluid motion as a result of thermally



# BACTERIAL HANDLING BY NON-UNIFORM ELECTRIC FIELDS




Fig. 9 – Fluid motion behavior at 10 kHz, for an interdigitated electrode array modeled by a 2-D FEM model. On the left, a few electrode domains are shown, while on the right-side a zoom of the neighborhood of an electrode is represented.


motion around the interdigitated microelectrodes consists of a couple of whirlpools rotating in opposite directions; while the left one spins counter-clockwise, the other one rotates clockwise. If there are particles in the solution and the applied voltage is high enough, those particles could show the tendency to concentrate at the center of the electrodes.

Figure 9 shows that particles are drawn down by fluid motion onto the electrode edge and then into the center, where they collect. This effect is dependent on frequency as shown in our experiments with bacterial cells. In view of the practical results, it is evident that other normally unconsidered effects (e.g. EHD and electroosmotic effects), which in addition to the primarily studied dielectrophoretic force also influence the particle motion, must be taken into account when working at low-frequency ranges. Indeed, such a motion actually

## CONCLUSIONS

This work was conceived to explore the possibilities of microsystems aimed at bioparticle handling based on common dielectrophoresis. To attain this goal, the main efforts were orientated towards the modeling and the design of interdigitated castellated microelectrode arrays that could adapt themselves to the electrohandling of both artificial and natural microparticles.

In order to achieve a good understanding of the related phenomena, modeling tasks begin with the description of the electrical model, which includes the particle modeling and the electrical field calculation. A 3-D finite element model was employed to know the electric field profile which is helpful in understanding and predicting the particle behavior. It is worth mentioning that the 3-D approach gives a bit more insight



From a technological point of view, a whole microsystem was designed and fabricated by means of microfabrication techniques. Such a microstructure includes a silicon substrate onto which electrodes of platinum were grown by lift-off technique to fulfill biocompatibility requirements.

From the experimental point of view, microstructures were tested utilizing *Escherichia coli* as bioparticle models. Bacterial cells moved towards the edges of castellated microelectrode arrays and were trapped there, as expected from the electric field calculations, being affected by positive dielectrophoresis when the applied signal was greater than 100 kHz. Such bacterial behavior is coherent with that explained by the single-shell particle model but it fails at lower frequencies where predicts the occurrence of negative dielectrophoresis.

As the frequency was decreased, bioparticles were observed to detach from points of maximum electric field strength and move towards the electrode center, clustering there. Nowadays this behavior is attributed to the fluid motion over the electrodes, which moves particles away from the electrode edges and into well-defined regions on top of the electrodes. In fact, numerical calculations of effects, i.e. fluid motion as a result of thermally induced medium inhomogeneities, demonstrated that at low frequencies the fluid motion around the electrodes, consisting of a couple of whirlpools rotating in opposite directions, can detach the particles from the electrode edges dragging them into the center, where they collect.

In view of the practical results, it is evident that other normally unconsidered effects (e.g. EHD and electroosmotic effects), which in addition to the primarily studied dielectrophoretic force also influence the particle motion, must be taken into account when designing and studying DEP-based handling microsystems.

To conclude one can say that the experiments illustrate the convenience of using castellated-platinum microelectrodes to microhandling *E. coli*, opening the possibility for using them with other kind of bioparticles. Furthermore, viscous drag particle motion as a result of


## RESUMO

Este artigo descreve a modelagem e teste experimental de uma rede de microeletrodos em cremalheira cujo objetivo é o manuseio de células biológicas, com base em dielectroforese comum. O microsistema proposto foi desenvolvido empregando eletrodos de platina depositados por técnicas de ‘lift-off’, micro-usinagem em silício e litografia com foto-resina. Uma vez fabricado o microdispositivo, este foi testado utilizando a *Escherichia coli* como modelo de biopartículas. Dielectroforese positiva pode ser observada com as células selecionadas para frequências acima de 100kHz, e efeitos eletro-hidrodinâmicos foram observados como o fenômeno dominante para menores frequências. Como resultado, a dielectroforese negativa não pode ser observada pois sua ocorrência se sobrepõe a efeitos eletro-hidrodinâmicos; i.e. a força de arraste viscoso atuando sobre as partículas é superior à força dielectroforética para frequências em que a dielectroforese negativa deveria ocorrer. Os experimentos ilustram a conveniência deste tipo de microdispositivo para o micromanuseio de objetos biológicos, abrindo a possibilidade de uso destas micro-redes com outras partículas biológicas. Além disto, o movimento líquido como resultado dos efeitos eletro-hidrodinâmicos deve ser levado em conta ao se desenhar micromanipuladores de partículas biológicas, e pode ser utilizado como mecanismo para limpar as superfícies dos eletrodos, que é um dos problemas mais importantes relacionados a este tipo de dispositivo.

**Palavras-chave:** dielectroforese, efeitos eletro-hidrodinâmicos, ‘biochips’, manuseio de bactérias.


## REFERENCES

- ARNOLD WM, SCHWAN HP AND ZIMMERMANN U. 1987. Surface conductance and other properties of latex particles measured by electrorotation. *J Phys Chem* 91: 5093–5098.
- BROWN AP. 1996. Dielectrophoretic investigations of bacterial cells. Doctoral thesis, University of York, York, United Kingdom.
- BURT JPH, PETHIG R, GASCOYNE PRC AND BECKER FF. 1990. Dielectrophoretic characterization of Friend murine erythroleukaemic cells as a measure of induced differentiation. *Biochim Biophys Acta* 1034: 93–101.
- CHENG J, SHELDON E, WU L, URIBE A, GERRUE L, CAR-



# BACTERIAL HANDLING BY NON-UNIFORM ELECTRIC FIELDS

- CHOI JW, BHANSALI S, AHN C AND THURMAN H. 1999. A new magnetic bead-based filterless bio-separator for integrated bio-molecule detection systems. The 13th European Conference on Solid-State Transducers EUROSENSORS XIII, The Hague, The Netherlands, September 12-15, p. 363–364.
- FEELEY CM AND POHL HA. 1981. The influence of resistivity and permittivity on the motion of uncharged solid particles in non-uniform electric fields (the ‘dielectrophoretic effect’). *J Phys D Appl Phys* 14: 2129–2138.
- FERNÁNDEZ F. 2000. Design, assembly and testing of microsystems for dielectrophoresis-based bioparticle electrohandling. Doctoral thesis, Department of Electronics, University of Barcelona, Barcelona, Spain.
- FERNÁNDEZ FH, DUARTE JE AND SAMITIER J. 2002. Micro-belt conveyor of latex microspheres. 17th Symposium on Microelectronics Technology and Devices, SB-Micro, Porto Alegre, Brazil, September 9-14, p. 133–140.
- FIGEYS D AND PINTO D. 2000. Lab-on-a-chip: A revolution in biological and medical and medical sciences. *Anal Chem* 72: 330a–335a.
- FUHR G AND SHIRLEY SG. 1998. Biological application of microstructures. *Top current chem* 194: 83–116.
- FUHR G AND WAGNER B. 1994. Electric field mediated cell manipulation in highly conductive culture media. *Microsystem technologies ‘94/4th International Conference on Micro, Electro, Opto, Mechanical Systems and Components*, Berlin, Germany, p. 407–416.
- FUHR G, SCHNELLE T, MÜLLER T, HITZLER H, MONAJEMBASHI S AND GREULICH K. 1998a. Force measurements of optical tweezers in electro-optical cages. *Appl Phys A* 67: 385–390.
- FUHR G, MÜLLER T, BAUKLOH V AND LUCAS K. 1998b. High-frequency electric field trapping of individual human spermatozoa. *Hum Reprod* 13: 136–141.
- GASCOYNE PRC, BECKER FF AND WANG XB. 1995. Numerical analysis of the influence of experimental conditions on the accuracy of dielectric parameters derived from electrorotation measurements. *Bioelectrochem Bioenerg* 36: 115–125.
- GASCOYNE PRC, WANG XB, HUANG Y AND BECKER FF. 1997. Dielectrophoretic separation of cancer cells from blood. *IEEE Trans Ind Appl* 33: 670–678.
- GREEN NG AND MORGAN H. 1998. Separation of submicrometre particles using a combination of dielectrophoretokinetics: a survey of sub-micrometre particle dynamics. *J Phys D Appl Phys* 33: 632–641.
- GREEN NG, RAMOS A AND GONZÁLEZ A. 2000b. Electric field induced fluid flow motion on microelectrodes: effect of illumination. *J Phys D Appl Phys* 33: L1–L4.
- HOETTGES K, HUGHES M, COTTON A, HOPKINS M AND MCDONELL M. 2003. Optimizing particle collection enhanced surface-based biosensors. *IEEE Eng Med Biol Mag* 22: 68–74.
- HUGHES MP AND MORGAN H. 1999. Measurement of cellular bacterial thrust by negative dielectrophoresis. *IEEE Trans Tech Prog* 15: 245–249.
- KALER K AND POHL HA. 1983. Dynamic dielectrophoretic levitation of living individual cells. *IEEE Trans Ind Appl* IA-19: 1089–1093.
- KOHNKE P. 1995. ANSYS Theory Reference Manual, Version 5.5. Swanson Analysis Systems Inc.
- LEE SW AND TAI YC. 1999. A micro cell lysis device. *IEEE Trans Actuators A* 73: 74–79.
- LEE SW, CHOI JH AND KIM YK. 1995. Design of a microcellular cell fusion device. The 8th International Conference on Solid-State Sensors and Actuators, and Eurosen- sors’95 – EUROSENSORS IX, Stockholm, Sweden, June 25-29, p. 377–380.
- MARKX GH, HUANG Y, ZHOU XF AND PETHIG R. 1994. Dielectrophoretic characterization and separation of micro-organisms. *Microbiology* 140: 585–591.
- NISHIOKA M, KATSURA S, HIRANO K AND MIZUTANI T. 1997. Evaluations of cell characteristics by step-wise rotational rotation using optoelectrostatic micromanipulation. *IEEE Trans Ind Appl* 33: 1381–1388.
- PETHIG R, HUANG Y, WANG XB AND BURT JPH. 1995. Positive and negative dielectrophoretic collection of colloidal particles using interdigitated castellated microelectrodes. *J Phys D Appl Phys* 24: 881–888.
- PETHIG R, BURT JPH, PARTON A, RIZVI N, TALBOT JH AND TAME JA. 1998. Development of biofactorization chip technology using excimer laser micromachining. *Micromech Microeng* 14: 57–63.
- POHL HA. 1951. The motion and precipitation of suspensions in divergent electric fields. *J Appl Phys* 22: 869–874.
- PORRAS YM, PEDRAZA OA, FERNÁNDEZ F, FERNÁNDEZ F AND DUARTE JE. 2004. Manipulación de protozoos por campos eléctricos. *Rev Col Biotech* 6: 79–84.



- REIMER K, KÖHLER C, LISEC T, SCHNAKENBERG U, FUHR G, HINTSCHE R AND WAGNER B. 1995. Fabrication of electrode arrays in the quarter micron regime for biotechnological applications. *Sens Actuators A* 46-47: 66–70.
- ROSENTHAL A AND VOLDMAN J. 2005. Dielectrophoretic traps for single-particle patterning. *Biophys J* 88: 2193–2205.
- SATO K, KAWAMURA Y, TANAKA S, UCHIDA K AND KOHIDA H. 1990. Individual and mass operation of biological cells using micromechanical silicon devices. *Sens Actuators A* 21-A23: 948–953.
- SCHNELLE T, MÜLLER T, VOIGT A, REIMER K, WAGNER B AND FUHR G. 1996. Adhesion-Inhibited surfaces. Coated and uncoated interdigitated electrode arrays in the micrometer and submicrometer range. *Langmuir* 12: 801–809.
- STEPHENS M, TALARY M, PETHIG R, BURNETT A AND MILLS K. 1996. The dielectrophoresis enrichment of CD34+ cells from peripheral blood stem cell harvests. *Bone Marrow Transpl* 18: 777–782.
- VELEV O AND KALER E. 1999. In situ assembly of colloidal particles into miniaturized biosensors. *Langmuir* 15: 3693–3698.
- WANG XB, HUANG Y, BURT JPH, MARKX GH AND PETHIG R. 1993. Selective dielectrophoretic confinement of bioparticles in potential energy wells. *J Phys D Appl Phys* 26: 1278–1285.
- WANG XB, HUANG Y, GASCOYNE PRC AND BECKER FF. 1997. Dielectrophoretic manipulation of particles. *IEEE Trans Ind Appl* 33: 660–669.
- WEBSTER JR AND MASTRANGELO CH. 1997. Large-volume integrated capillary electrophoresis stage fabricated using micromachining of plastics on silicon substrates. *International Conference on Solid-State Sensors and Actuators TRANSDUCERS'97*, Chicago, Illinois, USA, June 16-19, 503–506.