



Jornal Brasileiro de Patologia e Medicina Laboratorial

ISSN: 1676-2444

jbpm@sbpc.org.br, adagmar.andriolo@gmail.com

Sociedade Brasileira de Patologia Clínica/Medicina Laboratorial

Jenne Mimica, Marcelo; Dalla Valle Martino, Marínes; Pasternak, Jacyr  
MALDI-TOF MS in the clinical microbiology laboratory  
Jornal Brasileiro de Patologia e Medicina Laboratorial, vol. 49, núm. 4, agosto, 2013, pp.  
256-259  
Sociedade Brasileira de Patologia Clínica/Medicina Laboratorial  
Rio de Janeiro, Brasil

Available in: <http://www.redalyc.org/articulo.oa?id=393541977004>

- How to cite
- Complete issue
- More information about this article
- Journal's homepage in redalyc.org

redalyc.org

Scientific Information System  
Network of Scientific Journals from Latin America, the Caribbean, Spain and Portugal  
Non-profit academic project, developed under the open access initiative

# MALDI-TOF MS in the clinical microbiology laboratory

## *MALDI-TOF MS no laboratório de microbiologia clínica*

Marcelo Jenne Mimica<sup>1</sup>; Marines Dalla Valle Martino<sup>2</sup>; Jacyr Pasternak<sup>3</sup>

### ABSTRACT

Traditional methods for microbial identification are often very laborious and time consuming. A new mass spectrometry based technique, matrix-assisted laser desorption ionization-time of flight (MALDI-TOF), has been described as a rapid, practical and low-cost method for this purpose. In this article, primary and possible future applications of this tool are briefly discussed.

**Key words:** MALDI-TOF; microbial diagnostics; microbial identification; mass spectrometry.

### INTRODUCTION

Phenotypic methods for bacterial identification, including the Gram group, growth characteristics in culture and biochemical tests, remain the tests of choice for performing microbiological analysis in most clinical laboratories worldwide. Although these are traditional methods, they provide answers to the clinician that are often time consuming, reaching several days in the case of fastidious microorganisms<sup>(3)</sup>, hence the need for faster and accurate diagnostic tools. Some molecular techniques such as real-time polymerase chain reaction (RT-PCR), sequencing and microarrays have been successfully deployed in identification, but they are not widespread yet due to some practical limitations.

A mass spectrometry technique, matrix-assisted laser desorption ionization-time of flight (MALDI-TOF), has been used in the analysis of different biomolecules for decades, and its application in routine microbiological diagnosis seems to be promising<sup>(1, 21)</sup>. After its refinement in recent years, this technique has been described as an effective tool in the identification of Gram-positive, enterobacteriaceae, non-fermentative, anaerobic, mycobacteria and fungi<sup>(6, 10, 17, 24, 26, 30)</sup>. The method is based on protein analysis (the analysis of nucleic acids is also feasible, but rather cumbersome for this purpose) and comparison with available profiles stored in a database<sup>(14)</sup>.

Most studies with this method included strains that had previously been isolated in culture. Nevertheless, some reports admit the possibility of diagnosis directly from clinical specimens, especially blood samples, which would represent a breakthrough in terms of even faster results<sup>(3)</sup>.

### METHOD

There are two commercially available systems that have been incorporated in the microbiology area by traditional companies. The first one is MS Vitek (bioMérieux), formerly Shimadzu. The other one is Microflex Biotyper (Bruker), commercialized by Becton Dickinson (BD). Their most distinctive features are related to the operating system, the routine management benefits, size, use of disposable or non-disposable plates, the cost benefits of tests and use of plates, the control quality frequency and noise<sup>(27)</sup>.

The process initializes with the ionization of the sample to be studied, which is accomplished by mixing it with a matrix (alpha-cyano-4-hydroxycinnamic acid) and subsequently exposing it to laser pulses. Thus, the ionized molecules are accelerated by an electrostatic field, passing through a field-free area afterwards and, finally, reaching the mass spectrometry detector. Molecules of different masses and charges “fly” (hence the term time of flight)

First submission on 11/12/12; last submission on 11/12/12; accepted for publication on 13/12/12; published on 20/08/13

1. Doctor in Pediatrics by Santa Casa de São Paulo-Medical School (SCSP-MS); assistant professor of Microbiology at SCSP-MS.

2. Doctor in Pediatrics by SCSP-MS; adjunct professor of Microbiology at SCSP-MS; coordinator at the Department of Microbiology at Sociedade Beneficente Israelita Brasileira do Hospital Israelita Albert Einstein (SBIBHAE)-Clinical Laboratory.

3. Doctor in medicine by Universidade Estadual de Campinas (UNICAMP); doctor at the Department of Microbiology at SBIBHAE Clinical Laboratory; president of the Committee of Hospital Infection Control at HAE.

at different speeds. The results are compared with a database, and depending on the spectrum, it is possible to classify the microorganism into genus or species<sup>(14)</sup>.

## IDENTIFICATION OF ISOLATED BACTERIA

Seng *et al.*<sup>(32)</sup> prospectively analyzed 1,660 strains of bacteria identified in parallel through MALDI-TOF and traditional phenotypic methods. 84.1% were accurately identified at the species level and 11.3% at the genus level. According to the authors, no identification (2.8%) and misidentification (1.7%) were due to flaws in the database system. Van Veen *et al.*<sup>(36)</sup> also analyzed 980 strains prospectively, detecting a 6.8% identification of genus and 92% identification of species. In a retrospective study, Eigner *et al.*<sup>(13)</sup> studied 1,116 bacterial strains previously identified in clinical routine. MALDI-TOF consistency was 95.2%. Following a prospective study, Bizzini *et al.*<sup>(4)</sup> obtained correct identification of 95.1 % of 1,278 strains at the species level and 3% at the genus level. Other studies have yielded similar results<sup>(7, 29)</sup>.

Comparing the two available systems in a study that comprised 1,129 samples, there was 92.7% consistency at the species level versus 93.2%, accurately identified by Biotyper or Vitek MS, with no significant statistical difference<sup>(27)</sup>.

As for the test costs, these studies demonstrated that MALDI-TOF was more cost effective in comparison with traditional methods. Even when we take into account the price of the equipment (with depreciation diluted in the first years of use), it is estimated that the cost reaches less than 25% of the conventional identification<sup>(7, 29, 32)</sup>. Apart from the equipment purchase, the annual maintenance cost becomes more significant rather than the inputs. Results are available on average after only a few minutes, representing a significant difference in comparison with the time required for the results of conventional methods<sup>(29, 32)</sup>.

In general, although adjustments and improvements are required, mainly for some bacteria such as viridans group streptococci, *Shigella* and anaerobes<sup>(3)</sup>, the accuracy of this identification method seems to be considerable, with higher performance and cost effectiveness per identified strain.

## IDENTIFICATION OF ISOLATED FUNGI

This technique has been deployed in the characterization of various species of yeast and filamentous fungi. Studies with filamentous fungi are less numerous, probably due to the difficulties related to phenotypic heterogeneity and lack of standardization in protein extraction protocols. Even so, some investigations have

demonstrated the success of MALDI-TOF in identifying *Penicillium*, *Aspergillus*, *Fusarium* and some dermatophytes<sup>(15, 18, 19, 24)</sup>. For yeasts, mainly the genus *Candida*, the high identification accuracy has been established<sup>(26, 33, 36)</sup>, and the available databases include reference spectrum of all species commonly found in laboratories.

## OTHER APPLICATIONS

After successfully using the tool for the identification of microorganisms already isolated, the major challenge is to perform it directly from clinical samples. Although the reports on microorganism detection by MALDI-TOF directly from positive blood cultures are still relatively scarce<sup>(8, 16, 20, 22, 31, 34)</sup>, the accuracy seems to reach more than 80% when there is an inoculum higher than 10<sup>7</sup> organisms/ml. A preliminary stage of pre-processing extraction and purification should be carried out, which aims at microorganism concentration as well as separation of proteins from leukocytes and erythrocytes that may be interfering factors. To do so, “in house” method or commercial kit (Sepsityper) may be deployed<sup>(5, 34)</sup>.

Some hindrances that must be overcome with the improvement of the method are the possibility of failure in the differentiation between pneumococci and viridans streptococci and the difficulty in identifying two or more organisms in the sample<sup>(11)</sup>, a problem already solved depending on the system and version of the software in use. Furthermore, other applications for this method have been described in the literature, including the assessment of antimicrobial resistance<sup>(12, 23, 25)</sup> and the detection of virulence factors<sup>(2)</sup>, in addition to the use for classification and grouping with taxonomic or epidemiological purposes<sup>(28)</sup>. With the advance of this tool and methods of sample refinement prior to the use of MALDI-TOF, other applications as well as the detection of pathogens directly from clinical samples tend to be facilitated<sup>(3, 9, 35)</sup>.

## CONCLUSION

Mass spectrometry by MALDI-TOF could revolutionize diagnosis in clinical microbiology, providing accurate, inexpensive and quick identification of microorganisms. Its exact role in the routine laboratory tests remains to be defined, but the technique has the potential to be a major part of microbiological routine, along with traditional laboratory tests. New applications of the method, including the detection of pathogens directly from clinical samples, as well as assessment of virulence factors, antimicrobial susceptibility and even bacterial typing, make MALDI-TOF even more promising. It is expected that the more update the database is, the more accurate the bacterial identification will be.

## RESUMO

Os métodos tradicionais para identificação de microrganismos no laboratório clínico muitas vezes são trabalhosos e demorados. Uma nova metodologia, com base em espectrometria de massas, a matrix-assisted laser desorption ionization-time of flight (MALDI-TOF), é extremamente promissora para utilização na rotina microbiológica, sendo rápida, prática e pouco custosa. Neste artigo, são expostas, de forma breve, as principais aplicações atuais do método, assim como as perspectivas futuras.

**Unitermos:** MALDI-TOF; diagnóstico microbiológico; identificação microbiana; espectrometria de massas.

## REFERENCES

1. ANHALT, J. P.; FENSELAU, C. Identification of bacteria using mass spectrometry. *Anal Chem*, v. 47, p. 219-25, 1975.
2. BITTAR, F. *et al.* MALDI-TOF-MS for rapid detection of staphylococcal Pantón-Valentine leukocidin. *Int J Antimicrob Agents*, v. 34, p. 467-70, 2009.
3. BIZZINI, A.; GREUB, G. Matrix-assisted laser desorption ionization time-of-flight mass spectrometry, a revolution in clinical microbial identification. *Clin Microbiol Infect*, v. 16, p. 1614-9, 2010.
4. BIZZINI, A. *et al.* Performance of matrix-assisted laser desorption ionization-time of flight mass spectrometry for identification of bacterial strains routinely isolated in a clinical microbiology laboratory. *J Clin Microbiol*, v. 48, p. 1549-54, 2010.
5. BUCHAN, B. W.; RIEBE, K. M.; LEDEBOER, N. A. Comparison of the MALDI biotyper system using sepsityper specimen processing to routine microbiological methods for identification of bacteria from positive blood culture bottles. *J Clin Microbiol*, v. 50, p. 346-52, 2012.
6. CARBONNELLE, E. *et al.* Rapid identification of *Staphylococci* isolated in clinical microbiology laboratories by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol*, v. 45, p. 2156-61, 2007.
7. CHERKAoui, A. *et al.* Comparison of two matrix-assisted laser desorption ionization-time of flight mass spectrometry methods with conventional phenotypic identification for routine identification of bacteria to the species level. *J Clin Microbiol*, v. 48, p. 1169-75, 2010.
8. CHRISTNER, M. *et al.* Rapid identification of bacteria from positive blood culture bottles by use of matrix-assisted laser desorption-ionization time of flight mass spectrometry fingerprinting. *J Clin Microbiol*, v. 48, p. 1584-91, 2010.
9. CROXATTO, A.; PRODHOM, G.; GREUB, G. Applications of MALDI-TOF mass spectrometry in clinical diagnostic microbiology. *FEMS Microbiol Rev*, v. 36, p. 380-407, 2012.
10. DEGAND, N. *et al.* Matrix-assisted laser desorption ionization-time of flight mass spectrometry for identification of nonfermenting gram-negative bacilli isolated from cystic fibrosis patients. *J Clin Microbiol*, v. 46, p. 3361-7, 2008.
11. DRANCOURT, M. Detection of microorganisms in blood specimens using matrix-assisted laser desorption ionization time-of-flight mass spectrometry: a review. *Clin Microbiol Infect*, v. 16, p. 1620-5, 2010.
12. EDWARDS-JONES, V. *et al.* Rapid discrimination between methicillin-sensitive and methicillin-resistant *Staphylococcus aureus* by intact cell mass spectrometry. *J Med Microbiol*, v. 49, p. 295-300, 2000.
13. EIGNER, U. *et al.* Performance of a matrix-assisted laser desorption ionization-time-of-flight mass spectrometry system for the identification of bacterial isolates in the clinical routine laboratory. *Clin Lab*, v. 55, p. 289-96, 2009.
14. EMONET, S. *et al.* Application and use of various mass spectrometry methods in clinical microbiology. *Clin Microbiol Infect*, v. 16, p. 1604-13, 2010.
15. ERHARD, M. *et al.* Identification of dermatophyte species causing onychomycosis and tinea pedis by MALDI-TOF mass spectrometry. *Exp Dermatol*, v. 17, p. 356-61, 2008.
16. FERRONI, A. *et al.* Real-time identification of bacteria and *Candida* species in positive blood culture broths by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol*, v. 48, p. 1542-8, 2010.
17. GROSSE-HERRENTHEY, A. *et al.* Challenging the problem of clostridial identification with matrix-assisted laser desorption and ionization-time-of-flight mass spectrometry (MALDI-TOF MS). *Anaerobe*, v. 14, p. 242-9, 2008.
18. HETTICK, J. M. *et al.* Discrimination of *Penicillium* isolates by matrix-assisted laser desorption/ionization time-of-flight mass spectrometry fingerprinting. *Rapid Commun Mass Spectrom*, v. 22, p. 2555-60, 2008.
19. HETTICK, J. M. *et al.* Discrimination of *Aspergillus* isolates at the species and strain level by matrix-assisted laser desorption/ionization time-of-flight mass spectrometry fingerprinting. *Anal Biochem*, v. 380, p. 276-81, 2008.
20. KALETA, E. J. *et al.* Comparative analysis of PCR-electrospray ionization/mass spectrometry (MS) and MALDI-TOF/MS for the identification of bacteria and yeast from positive blood culture bottles. *Clin Chem*, v. 57, p. 1057-67, 2011.
21. KARAS, M.; BACHMANN, D.; HILLENKAMP, F. Influence of the wavelength in high-irradiance ultraviolet laser desorption mass spectrometry of organic molecules. *Anal Chem*, v. 57, p. 2935-9, 1985.
22. LA SCOLA, B.; RAOULT, D. Direct identification of bacteria in positive blood culture bottles by matrix-assisted laser desorption ionization time-of-flight mass spectrometry. *PLoS One*, v. 4, e8041, 2009.
23. MAJCHERCZYK, P. A. *et al.* The discriminatory power of MALDI-TOF mass spectrometry to differentiate between isogenic teicoplanin-

susceptible and teicoplanin-resistant strains of methicillin-resistant *Staphylococcus aureus*. *FEMS Microbiol Lett*, v. 255, p. 233-9, 2006.

24. MARINACH-PATRICE, C. *et al.* Use of mass spectrometry to identify clinical *Fusarium* isolates. *Clin Microbiol Infect*, v. 15, p. 634-42, 2009.

25. MARINACH, C. *et al.* MALDI-TOF MS-based drug susceptibility testing of pathogens: the example of *Candida albicans* and fluconazole. *Proteomics*, v. 9, p. 4627-31, 2009.

26. MARKLEIN, G. *et al.* Matrix-assisted laser desorption ionization-time of flight mass spectrometry for fast and reliable identification of clinical yeast isolates. *J Clin Microbiol*, v. 47, p. 2912-7, 2009.

27. MARTINY, D. *et al.* Comparison of Microflex LT and Vitek MS systems for routine identification of bacteria by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol*, v. 50, p. 1313-25, 2012.

28. MURRAY, P. R. Matrix-assisted laser desorption ionization time-of-flight mass spectrometry: usefulness for taxonomy and epidemiology. *Clin Microbiol Infect*, v. 16, p. 1626-30, 2010.

29. NEVILLE, S. A. *et al.* Utility of matrix-assisted laser desorption ionization-time of flight mass spectrometry following introduction for routine laboratory bacterial identification. *J Clin Microbiol*, v. 49, p. 2980-4, 2011.

30. PIGNONE, M. *et al.* Identification of mycobacteria by matrix-assisted laser desorption ionization-time-of-flight mass spectrometry. *J Clin Microbiol*, v. 44, p. 1963-70, 2006.

31. PRODHOM, G. *et al.* Matrix-assisted laser desorption ionization-time of flight mass spectrometry for direct bacterial identification from positive blood culture pellets. *J Clin Microbiol*, v. 48, p. 1481-3, 2010.

32. SENG, P. *et al.* Ongoing revolution in bacteriology: routine identification of bacteria by matrix-assisted laser desorption ionization time-of-flight mass spectrometry. *Clin Infect Dis*, v. 49, p. 543-51, 2009.

33. SEYFARTH, F. *et al.* Identification of yeast isolated from dermatological patients by MALDI-TOF mass spectrometry. *Mycoses*, v. 55, p. 276-80, 2012.

34. STEVENSON, L. G.; DRAKE, S. K.; MURRAY, P. R. Rapid identification of bacteria in positive blood culture broths by matrix-assisted laser desorption ionization-time of flight mass spectrometry. *J Clin Microbiol*, v. 48, p. 444-7, 2010.

35. VAN BELKUM, A. *et al.* Biomedical mass spectrometry in today's and tomorrow's clinical microbiology laboratories. *J Clin Microbiol*, v. 50, p. 1513-7, 2012.

36. VAN VEEN, S. Q.; CLAAS, E. C.; KUIJPER, E. J. High-throughput identification of bacteria and yeast by matrix-assisted laser desorption ionization-time of flight mass spectrometry in conventional medical microbiology laboratories. *J Clin Microbiol*, v. 48, p. 900-7, 2010.

---

#### MAILING ADDRESS

Marcelo Jenne Mimica

Rua Cesário Motta Jr, 112; Departamento de Ciências Patológicas; Disciplina de Microbiologia; CEP: 01221-020; São Paulo-SP, Brazil; e-mail: mjmimica@hotmail.com..